

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052721\
 Data File : VX022201.D
 Acq On : 26 May 2021 16:42
 Operator : JC/MD
 Sample : M2526-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-117T-20210525

Quant Time: May 27 04:07:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	68720	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	143968	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	126616	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	50367	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	57390	54.14	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery = 108.28%			
35) Dibromofluoromethane	5.391	113	44225	48.44	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery = 96.88%			
50) Toluene-d8	8.653	98	175853	50.00	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery = 100.00%			
62) 4-Bromofluorobenzene	11.085	95	62731	48.26	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery = 96.52%			
Target Compounds						
						Qvalue
27) cis-1,2-Dichloroethene	4.507	96	1468	1.41	ug/l	97
44) Trichloroethene	7.129	130	18825	18.49	ug/l	96
95) Naphthalene	13.780	128	1825	0.46	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052721\
 Data File : VX022201.D
 Acq On : 26 May 2021 16:42
 Operator : JC/MD
 Sample : M2526-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-117T-20210525

Quant Time: May 27 04:07:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

