

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052721\
 Data File : VX022202.D
 Acq On : 26 May 2021 17:05
 Operator : JC/MD
 Sample : M2526-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-118.5N-20210525

Quant Time: May 27 04:07:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	67032	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	137777	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	122438	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	47290	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.971	65	54060	52.28	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	104.56%	
35) Dibromofluoromethane	5.397	113	41578	47.59	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	95.18%	
50) Toluene-d8	8.653	98	168973	50.21	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	100.42%	
62) 4-Bromofluorobenzene	11.085	95	58475	47.01	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	94.02%	
Target Compounds						
						Qvalue
16) Acetone	2.392	43	1527	3.39	ug/l #	87
95) Naphthalene	13.780	128	1451	0.39	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052721\
 Data File : VX022202.D
 Acq On : 26 May 2021 17:05
 Operator : JC/MD
 Sample : M2526-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
ClientSampleId :
 MH-118.5N-20210525

Quant Time: May 27 04:07:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

