

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052721\
 Data File : VX022205.D
 Acq On : 26 May 2021 18:16
 Operator : JC/MD
 Sample : M2526-15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-121.5T-20210525

Quant Time: May 27 04:08:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	65999	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	137422	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	118130	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	42881	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	53721	52.77	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.54%
35) Dibromofluoromethane	5.397	113	41943	48.13	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.26%
50) Toluene-d8	8.653	98	167525	49.90	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.80%
62) 4-Bromofluorobenzene	11.085	95	55211	44.50	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.00%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	1681	3.79	ug/l	88
27) cis-1,2-Dichloroethene	4.501	96	2675	2.67	ug/l	93
44) Trichloroethene	7.135	130	1792	1.84	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052721\
Data File : VX022205.D
Acq On : 26 May 2021 18:16
Operator : JC/MD
Sample : M2526-15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-121.5T-20210525

Quant Time: May 27 04:08:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

