

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX052721\  
 Data File : VX022209.D  
 Acq On : 26 May 2021 19:50  
 Operator : JC/MD  
 Sample : M2526-22 100X  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 23 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 DUP01-20210525

Quant Time: May 27 04:09:32 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X051921W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed May 19 14:23:10 2021  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	5.562	168	68077	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	145433	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	128020	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	50092	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	57456	54.71	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	109.42%
35) Dibromofluoromethane	5.397	113	44611	48.37	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.74%
50) Toluene-d8	8.653	98	177801	50.05	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.10%
62) 4-Bromofluorobenzene	11.085	95	61393	46.75	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.50%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX052721\  
 Data File : VX022209.D  
 Acq On : 26 May 2021 19:50  
 Operator : JC/MD  
 Sample : M2526-22 100X  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 23 Sample Multiplier: 1

**Instrument :**  
 MSVOA\_X  
**ClientSampled :**  
 DUP01-20210525

Quant Time: May 27 04:09:32 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X051921W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed May 19 14:23:10 2021  
 Response via : Initial Calibration

