

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052721\
 Data File : VX022210.D
 Acq On : 26 May 2021 20:14
 Operator : JC/MD
 Sample : M2526-23 100X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP02-20210525

Quant Time: May 27 04:09:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	65876	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	138418	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	116829	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	43498	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	54807	53.93	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	107.86%
35) Dibromofluoromethane	5.391	113	41757	47.57	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.14%
50) Toluene-d8	8.653	98	167789	49.62	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.24%
62) 4-Bromofluorobenzene	11.086	95	55691	44.56	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.12%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052721\
 Data File : VX022210.D
 Acq On : 26 May 2021 20:14
 Operator : JC/MD
 Sample : M2526-23 100X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
ClientSampleId :
 DUP02-20210525

Quant Time: May 27 04:09:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

