

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052721\
 Data File : VX022214.D
 Acq On : 26 May 2021 21:48
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 27 04:11:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	81	0.00
2 T	Dichlorodifluoromethane	0.608	0.600	1.3	76	0.00
3 P	Chloromethane	0.656	0.643	2.0	77	0.00
4 C	Vinyl Chloride	0.694	0.687	1.0#	77	0.00
5 T	Bromomethane	0.404	0.358	11.4	75	-0.01
6 T	Chloroethane	0.413	0.451	-9.2	83	0.00
7 T	Trichlorofluoromethane	0.998	1.038	-4.0	82	0.00
8 T	Diethyl Ether	0.429	0.428	0.2	81	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.592	0.580	2.0	78	0.00
10 T	Methyl Iodide	0.637	0.684	-7.4	82	0.00
11 T	Tert butyl alcohol	0.176	0.211	-19.9	99	0.00
12 CM	1,1-Dichloroethene	0.573	0.611	-6.6#	84	0.00
13 T	Acrolein	0.117	0.110	6.0	82	0.00
14 T	Allyl chloride	1.053	1.113	-5.7	83	0.00
15 T	Acrylonitrile	0.347	0.416	-19.9	94	0.00
16 T	Acetone	0.336	0.373	-11.0	93	0.00
17 T	Carbon Disulfide	1.501	1.327	11.6	74	0.00
18 T	Methyl Acetate	0.893	1.069	-19.7	98	0.00
19 T	Methyl tert-butyl Ether	1.963	2.199	-12.0	88	0.00
20 T	Methylene Chloride	0.746	0.750	-0.5	87	0.00
21 T	trans-1,2-Dichloroethene	0.623	0.667	-7.1	83	0.00
22 T	Diisopropyl ether	2.027	2.261	-11.5	89	0.00
23 T	Vinyl Acetate	1.952	2.199	-12.7	88	0.00
24 P	1,1-Dichloroethane	1.190	1.303	-9.5	87	0.00
25 T	2-Butanone	0.543	0.650	-19.7	94	0.00
26 T	2,2-Dichloropropane	1.056	0.883	16.4	67	0.00
27 T	cis-1,2-Dichloroethene	0.758	0.835	-10.2	85	0.00
28 T	Bromochloromethane	0.577	0.612	-6.1	88	0.00
29 T	Tetrahydrofuran	0.323	0.375	-16.1	94	0.00
30 C	Chloroform	1.229	1.356	-10.3#	88	0.00
31 T	Cyclohexane	0.969	0.963	0.6	80	-0.01
32 T	1,1,1-Trichloroethane	1.050	1.126	-7.2	85	0.00
33 S	1,2-Dichloroethane-d4	0.771	0.820	-6.4	94	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	85	0.00
35 S	Dibromofluoromethane	0.317	0.304	4.1	87	0.00
36 T	1,1-Dichloropropene	0.462	0.464	-0.4	83	0.00
37 T	Ethyl Acetate	0.567	0.607	-7.1	92	0.00
38 T	Carbon Tetrachloride	0.452	0.462	-2.2	81	0.00
39 T	Methylcyclohexane	0.524	0.488	6.9	79	0.00
40 TM	Benzene	1.423	1.470	-3.3	86	0.00
41 T	Methacrylonitrile	0.303	0.321	-5.9	89	0.00
42 TM	1,2-Dichloroethane	0.498	0.528	-6.0	88	0.00
43 T	Isopropyl Acetate	0.927	1.005	-8.4	91	0.00
44 TM	Trichloroethene	0.354	0.359	-1.4	84	0.00
45 C	1,2-Dichloropropane	0.363	0.396	-9.1#	87	0.00
46 T	Dibromomethane	0.246	0.266	-8.1	87	0.00
47 T	Bromodichloromethane	0.493	0.520	-5.5	84	0.00
48 T	Methyl methacrylate	0.427	0.471	-10.3	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052721\
 Data File : VX022214.D
 Acq On : 26 May 2021 21:48
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 27 04:11:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.010	-25.0	98	-0.02
50 S	Toluene-d8	1.221	1.201	1.6	90	0.00
51 T	4-Methyl-2-Pentanone	0.555	0.645	-16.2	94	0.00
52 CM	Toluene	0.901	0.935	-3.8#	86	0.00
53 T	t-1,3-Dichloropropene	0.569	0.572	-0.5	82	0.00
54 T	cis-1,3-Dichloropropene	0.611	0.617	-1.0	83	0.00
55 T	1,1,2-Trichloroethane	0.366	0.393	-7.4	88	0.00
56 T	Ethyl methacrylate	0.591	0.660	-11.7	89	0.00
57 T	1,3-Dichloropropane	0.624	0.681	-9.1	90	0.00
58 T	2-Chloroethyl Vinyl ether	0.320	0.329	-2.8	87	0.00
59 T	2-Hexanone	0.418	0.494	-18.2	95	0.00
60 T	Dibromochloromethane	0.356	0.378	-6.2	84	0.00
61 T	1,2-Dibromoethane	0.364	0.405	-11.3	89	0.00
62 S	4-Bromofluorobenzene	0.451	0.456	-1.1	92	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	85	0.00
64 T	Tetrachloroethene	0.328	0.309	5.8	78	0.00
65 PM	Chlorobenzene	1.041	1.071	-2.9	85	0.00
66 T	1,1,1,2-Tetrachloroethane	0.362	0.386	-6.6	86	0.00
67 C	Ethyl Benzene	1.915	1.946	-1.6#	85	0.00
68 T	m/p-Xylenes	0.724	0.729	-0.7	85	0.00
69 T	o-Xylene	0.704	0.737	-4.7	88	0.00
70 T	Styrene	1.153	1.230	-6.7	86	0.00
71 P	Bromoform	0.250	0.256	-2.4	79	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	86	0.00
73 T	Isopropylbenzene	4.232	4.215	0.4	86	0.00
74 T	N-amyl acetate	2.009	2.211	-10.1	93	0.00
75 P	1,1,2,2-Tetrachloroethane	1.403	1.551	-10.5	95	0.00
76 T	1,2,3-Trichloropropane	1.234	1.366	-10.7	92	0.00
77 T	Bromobenzene	0.897	0.897	0.0	85	0.00
78 T	n-propylbenzene	4.957	4.953	0.1	85	0.00
79 T	2-Chlorotoluene	3.006	2.982	0.8	87	0.00
80 T	1,3,5-Trimethylbenzene	3.507	3.513	-0.2	85	0.00
81 T	trans-1,4-Dichloro-2-butene	0.521	0.493	5.4	78	0.00
82 T	4-Chlorotoluene	3.464	3.403	1.8	86	0.00
83 T	tert-Butylbenzene	3.364	3.407	-1.3	86	0.00
84 T	1,2,4-Trimethylbenzene	3.530	3.579	-1.4	85	0.00
85 T	sec-Butylbenzene	4.275	4.329	-1.3	85	0.00
86 T	p-Isopropyltoluene	3.610	3.630	-0.6	85	0.00
87 T	1,3-Dichlorobenzene	1.730	1.751	-1.2	87	0.00
88 T	1,4-Dichlorobenzene	1.788	1.720	3.8	85	0.00
89 T	n-Butylbenzene	3.340	3.346	-0.2	83	0.00
90 T	Hexachloroethane	0.597	0.602	-0.8	82	0.00
91 T	1,2-Dichlorobenzene	1.662	1.746	-5.1	89	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.300	0.334	-11.3	94	0.00
93 T	1,2,4-Trichlorobenzene	1.003	1.012	-0.9	87	0.00
94 T	Hexachlorobutadiene	0.391	0.365	6.6	79	0.00
95 T	Naphthalene	3.910	4.319	-10.5	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052721\
Data File : VX022214.D
Acq On : 26 May 2021 21:48
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: May 27 04:11:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.011	1.057	-4.5	89	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6