

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052721\
 Data File : VX022189.D
 Acq On : 26 May 2021 11:54
 Operator : JC/MD
 Sample : VX0527MBL01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0527MBL01

Quant Time: May 26 13:11:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	67848	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	141453	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	121983	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	46911	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	54837	52.40	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.80%
35) Dibromofluoromethane	5.391	113	43970	49.02	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.04%
50) Toluene-d8	8.653	98	172529	49.93	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.86%
62) 4-Bromofluorobenzene	11.085	95	58517	45.82	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.64%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052721\
Data File : VX022189.D
Acq On : 26 May 2021 11:54
Operator : JC/MD
Sample : VX0527MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0527MBL01

Quant Time: May 26 13:11:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 5.562 min Scan# 734
Delta R.T. 0.000 min
Lab File: VX022189.D
Acq: 26 May 2021 11:54




Instrument :
MSVOA_X
ClientSampleId :
VX0527MBL01


Tgt Ion:168 Resp: 67848
Ion Ratio Lower Upper
168 100
99 68.3 45.8 68.6



#33
1,2-Dichloroethane-d4
Concen: 52.40 ug/l
RT: 5.958 min Scan# 799
Delta R.T. -0.006 min
Lab File: VX022189.D
Acq: 26 May 2021 11:54

Tgt Ion: 65 Resp: 54837
Ion Ratio Lower Upper
65 100
67 52.4 0.0 105.4









#34
1,4-Difluorobenzene
Concen: 50.00 ug/l
RT: 6.769 min Scan# 932
Delta R.T. 0.000 min
Lab File: VX022189.D
Acq: 26 May 2021 11:54




Instrument :
MSVOA_X
ClientSampleId :
VX0527MBL01


Tgt Ion:114 Resp: 141453
Ion Ratio Lower Upper
114 100
63 20.3 0.0 40.8
88 16.6 0.0 32.8



#35
Dibromofluoromethane
Concen: 49.02 ug/l
RT: 5.391 min Scan# 706
Delta R.T. 0.000 min
Lab File: VX022189.D
Acq: 26 May 2021 11:54

Tgt Ion:113 Resp: 43970
Ion Ratio Lower Upper
113 100
111 102.8 81.9 122.9
192 16.9 16.7 25.1









#50
Toluene-d8
Concen: 49.93 ug/l
RT: 8.653 min Scan# 1241
Delta R.T. -0.000 min
Lab File: VX022189.D
Acq: 26 May 2021 11:54




Instrument :
MSVOA_X
ClientSampleId :
VX0527MBL01


Tgt Ion: 98 Resp: 172529
Ion Ratio Lower Upper
98 100
100 66.9 51.9 77.9



#62
4-Bromofluorobenzene
Concen: 45.82 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. 0.000 min
Lab File: VX022189.D
Acq: 26 May 2021 11:54

Tgt Ion: 95 Resp: 58517
Ion Ratio Lower Upper
95 100
174 64.4 0.0 154.6
176 61.9 0.0 155.8









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.006 min
Lab File: VX022189.D
Acq: 26 May 2021 11:54

Instrument :
MSVOA_X
ClientSampleId :
VX0527MBL01

Tgt Ion:117 Resp: 121983
Ion Ratio Lower Upper
117 100
82 59.0 47.4 71.2
119 30.0 26.6 39.8



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX022189.D
Acq: 26 May 2021 11:54

Tgt Ion:152 Resp: 46911
Ion Ratio Lower Upper
152 100
115 66.4 41.1 123.3
150 160.1 0.0 349.0

