



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052721\
 Data File : VX022215.D
 Acq On : 26 May 2021 22:35
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X051321W.M
 Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 Last Update : Thu May 13 06:11:35 2021



AutoFind: Scans 1639, 1640, 1641; Background Corrected with Scan 1634

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.1	4417	PASS
75	95	30	60	55.3	12765	PASS
95	95	100	100	100.0	23069	PASS
96	95	5	9	7.5	1735	PASS
173	174	0.00	2	1.0	162	PASS
174	95	50	100	67.7	15613	PASS
175	174	5	9	7.2	1120	PASS
176	174	95	101	99.4	15514	PASS
177	176	5	9	6.8	1056	PASS