

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX052722\  
 Data File : VX029025.D  
 Acq On : 27 May 2022 15:37  
 Operator : JC/MD  
 Sample : N3028-02  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 11 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 TP-1-1-5FT

Quant Time: May 27 16:41:35 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X051222W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu May 12 14:58:53 2022  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units  | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|--------|----------|
| -----                       |        |       |          |          |        |          |
| Internal Standards          |        |       |          |          |        |          |
| 1) Pentafluorobenzene       | 5.556  | 168   | 353826   | 50.000   | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.763  | 114   | 629595   | 50.000   | ug/l   | 0.00     |
| 63) Chlorobenzene-d5        | 10.055 | 117   | 611033   | 50.000   | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.024 | 152   | 295642   | 50.000   | ug/l   | 0.00     |
| System Monitoring Compounds |        |       |          |          |        |          |
| 33) 1,2-Dichloroethane-d4   | 5.958  | 65    | 227891   | 46.987   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 61 - 141 | Recovery | =      | 93.980%  |
| 35) Dibromofluoromethane    | 5.391  | 113   | 205431   | 49.801   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 69 - 133 | Recovery | =      | 99.600%  |
| 50) Toluene-d8              | 8.653  | 98    | 752596   | 48.392   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 65 - 126 | Recovery | =      | 96.780%  |
| 62) 4-Bromofluorobenzene    | 11.085 | 95    | 302196   | 52.024   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 58 - 135 | Recovery | =      | 104.040% |
| Target Compounds            |        |       |          |          |        |          |
|                             |        |       |          |          | Qvalue |          |
| 16) Acetone                 | 2.380  | 43    | 62545    | 30.047   | ug/l   | 95       |
| 20) Methylene Chloride      | 2.788  | 84    | 25788    | 5.905    | ug/l   | 97       |
| 43) Isopropyl Acetate       | 6.348  | 43    | 28306    | 2.796    | ug/l   | 97       |
| -----                       |        |       |          |          |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX052722\  
Data File : VX029025.D  
Acq On : 27 May 2022 15:37  
Operator : JC/MD  
Sample : N3028-02  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 11 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
TP-1-1-5FT

Quant Time: May 27 16:41:35 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X051222W.M  
Quant Title : SW846 8260  
QLast Update : Thu May 12 14:58:53 2022  
Response via : Initial Calibration

