

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052722\
 Data File : VX029036.D
 Acq On : 27 May 2022 19:54
 Operator : JC/MD
 Sample : N3087-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-08

Quant Time: May 27 22:12:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	345599	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	611999	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	592346	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	285922	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	212852	44.931	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	89.860%
35) Dibromofluoromethane	5.391	113	197270	49.198	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.400%
50) Toluene-d8	8.653	98	725669	48.002	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.000%
62) 4-Bromofluorobenzene	11.079	95	289801	51.324	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	102.640%
Target Compounds						
					Qvalue	
30) Chloroform	5.105	83	36460	4.727	ug/l	95
52) Toluene	8.726	92	3378	0.303	ug/l	94
64) Tetrachloroethene	9.275	164	1070	0.239	ug/l #	89
67) Ethyl Benzene	10.195	91	16975	0.754	ug/l	99
68) m/p-Xylenes	10.305	106	17387	1.983	ug/l	96
69) o-Xylene	10.646	106	2061	0.239	ug/l	91
84) 1,2,4-Trimethylbenzene	11.756	105	24996	1.380	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052722\
Data File : VX029036.D
Acq On : 27 May 2022 19:54
Operator : JC/MD
Sample : N3087-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-08

Quant Time: May 27 22:12:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

