

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052722\
 Data File : VX029037.D
 Acq On : 27 May 2022 20:17
 Operator : JC/MD
 Sample : N3087-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-07

Quant Time: May 27 22:12:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	342867	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	601129	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	582488	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	274081	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	214659	45.673	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	91.340%		
35) Dibromofluoromethane	5.391	113	194691	49.433	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	98.860%		
50) Toluene-d8	8.653	98	709965	47.813	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	95.620%		
62) 4-Bromofluorobenzene	11.079	95	279652	50.423	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	100.840%		
Target Compounds						
					Qvalue	
64) Tetrachloroethene	9.275	164	1215	0.276	ug/l #	83
67) Ethyl Benzene	10.201	91	4791	0.216	ug/l	97
68) m/p-Xylenes	10.305	106	5393	0.625	ug/l	98
84) 1,2,4-Trimethylbenzene	11.756	105	6766	0.390	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052722\
Data File : VX029037.D
Acq On : 27 May 2022 20:17
Operator : JC/MD
Sample : N3087-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-07

Quant Time: May 27 22:12:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

