

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052722\
 Data File : VX029038.D
 Acq On : 27 May 2022 20:40
 Operator : JC/MD
 Sample : N3087-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-01A

Quant Time: May 27 22:12:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	347626	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	611821	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	592775	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	281842	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	220964	46.371	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	92.740%
35) Dibromofluoromethane	5.391	113	200345	49.979	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.960%
50) Toluene-d8	8.653	98	721237	47.723	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.440%
62) 4-Bromofluorobenzene	11.085	95	291760	51.686	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	103.380%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	9599	4.694	ug/l	97
17) Carbon Disulfide	2.520	76	2834	0.304	ug/l #	87
39) Methylcyclohexane	7.391	83	2760	0.400	ug/l #	85
40) Benzene	6.037	78	9729	0.562	ug/l	99
64) Tetrachloroethene	9.275	164	2189	0.489	ug/l	93
68) m/p-Xylenes	10.305	106	3686	0.420	ug/l	97
84) 1,2,4-Trimethylbenzene	11.756	105	4991	0.279	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052722\
Data File : VX029038.D
Acq On : 27 May 2022 20:40
Operator : JC/MD
Sample : N3087-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-01A

Quant Time: May 27 22:12:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

