

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052722\
 Data File : VX029042.D
 Acq On : 27 May 2022 23:23
 Operator : JC/MD
 Sample : VX0527WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0527WBL02

Quant Time: May 28 08:35:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	343057	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	608322	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	574770	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	270326	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	217116	46.170	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	92.340%
35) Dibromofluoromethane	5.391	113	197674	49.597	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.200%
50) Toluene-d8	8.653	98	715567	47.620	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.240%
62) 4-Bromofluorobenzene	11.079	95	278443	49.611	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	99.220%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052722\
Data File : VX029042.D
Acq On : 27 May 2022 23:23
Operator : JC/MD
Sample : VX0527WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0527WBL02

Quant Time: May 28 08:35:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

