

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052722\
 Data File : VX029047.D
 Acq On : 28 May 2022 01:19
 Operator : JC/MD
 Sample : N3022-16
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RB06-20220525

Quant Time: May 28 08:36:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	350028	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	621177	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	592349	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	280221	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	226201	47.144	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.280%
35) Dibromofluoromethane	5.391	113	205083	50.391	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.780%
50) Toluene-d8	8.653	98	732452	47.735	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.480%
62) 4-Bromofluorobenzene	11.079	95	288236	50.293	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	100.580%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	22116	10.740	ug/l	98
20) Methylene Chloride	2.788	84	2842	0.658	ug/l #	93
25) 2-Butanone	4.580	43	4612	1.427	ug/l #	85
44) Trichloroethene	7.135	130	1109	0.233	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052722\
Data File : VX029047.D
Acq On : 28 May 2022 01:19
Operator : JC/MD
Sample : N3022-16
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RB06-20220525

Quant Time: May 28 08:36:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

