

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052722\
 Data File : VX029050.D
 Acq On : 28 May 2022 02:29
 Operator : JC/MD
 Sample : N3022-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 56BR-20220525

Quant Time: May 28 08:37:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	340650	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	608231	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	598502	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	287999	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	219320	46.968	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.940%
35) Dibromofluoromethane	5.391	113	201118	50.468	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.940%
50) Toluene-d8	8.653	98	720219	47.937	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.880%
62) 4-Bromofluorobenzene	11.079	95	289719	51.628	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	103.260%
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.166	85	59094	13.248	ug/l	100
3) Chloromethane	1.294	50	913	0.217	ug/l	99
4) Vinyl Chloride	1.374	62	347128	76.681	ug/l	99
9) 1,1,2-Trichlorotrifluo...	2.331	101	105976	27.903	ug/l	98
12) 1,1-Dichloroethene	2.325	96	37061	10.512	ug/l	94
21) trans-1,2-Dichloroethene	3.093	96	20599	5.187	ug/l	96
27) cis-1,2-Dichloroethene	4.495	96	6686812	1441.707	ug/l	85
29) Tetrahydrofuran	5.032	42	2233	1.092	ug/l	93
44) Trichloroethene	7.159	130	22517463	4826.558	ug/l	97
64) Tetrachloroethene	9.275	164	42524	9.409	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052722\
Data File : VX029050.D
Acq On : 28 May 2022 02:29
Operator : JC/MD
Sample : N3022-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
56BR-20220525

Quant Time: May 28 08:37:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

