

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052722\
 Data File : VX029054.D
 Acq On : 28 May 2022 04:02
 Operator : JC/MD
 Sample : N3022-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 22BR-20220525

Quant Time: May 28 08:38:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	335069	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	589900	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	557253	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	256087	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	216033	47.035	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.080%
35) Dibromofluoromethane	5.391	113	196175	50.758	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.520%
50) Toluene-d8	8.653	98	696191	47.777	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.560%
62) 4-Bromofluorobenzene	11.079	95	268031	49.247	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	98.500%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.166	85	966	0.220	ug/l	80
3) Chloromethane	1.294	50	965	0.234	ug/l #	82
4) Vinyl Chloride	1.373	62	3760	0.844	ug/l	97
16) Acetone	2.379	43	3405	1.727	ug/l	95
20) Methylene Chloride	2.800	84	1034	0.250	ug/l #	79
27) cis-1,2-Dichloroethene	4.495	96	45834	10.047	ug/l	87
30) Chloroform	5.098	83	2634	0.352	ug/l	95
44) Trichloroethene	7.128	130	36978	8.172	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052722\
Data File : VX029054.D
Acq On : 28 May 2022 04:02
Operator : JC/MD
Sample : N3022-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
22BR-20220525

Quant Time: May 28 08:38:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

