

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052722\
 Data File : VX029055.D
 Acq On : 28 May 2022 04:26
 Operator : JC/MD
 Sample : N3022-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 48BR-20220525

Quant Time: May 28 08:39:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	325973	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	582384	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	560482	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	263145	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	209037	46.782	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.560%
35) Dibromofluoromethane	5.391	113	191216	50.113	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.220%
50) Toluene-d8	8.653	98	684871	47.607	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.220%
62) 4-Bromofluorobenzene	11.085	95	268583	49.985	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	99.980%
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.374	62	3293	0.760	ug/l	93
9) 1,1,2-Trichlorotrifluo...	2.331	101	4934	1.358	ug/l	98
12) 1,1-Dichloroethene	2.319	96	11946	3.541	ug/l	96
21) trans-1,2-Dichloroethene	3.093	96	1238	0.326	ug/l #	87
24) 1,1-Dichloroethane	3.617	63	4214	0.627	ug/l	99
27) cis-1,2-Dichloroethene	4.489	96	67749	15.265	ug/l	86
44) Trichloroethene	7.129	130	2020795	452.376	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052722\
Data File : VX029055.D
Acq On : 28 May 2022 04:26
Operator : JC/MD
Sample : N3022-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
48BR-20220525

Quant Time: May 28 08:39:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

