

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052722\
 Data File : VX029056.D
 Acq On : 28 May 2022 04:49
 Operator : JC/MD
 Sample : N3022-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 08BR-20220525

Quant Time: May 28 08:39:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	326458	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	584092	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	560214	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	259251	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	213470	47.703	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.400%
35) Dibromofluoromethane	5.391	113	191115	49.940	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.880%
50) Toluene-d8	8.653	98	695701	48.219	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.440%
62) 4-Bromofluorobenzene	11.079	95	270287	50.155	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	100.320%
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.166	85	1253	0.293	ug/l	91
9) 1,1,2-Trichlorotrifluo...	2.330	101	21318	5.857	ug/l	98
12) 1,1-Dichloroethene	2.324	96	1736	0.514	ug/l	94
16) Acetone	2.385	43	4941	2.573	ug/l #	85
21) trans-1,2-Dichloroethene	3.093	96	1542	0.405	ug/l	93
24) 1,1-Dichloroethane	3.617	63	3547	0.527	ug/l	97
27) cis-1,2-Dichloroethene	4.495	96	48262	10.858	ug/l	87
44) Trichloroethene	7.128	130	790857	176.524	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052722\
Data File : VX029056.D
Acq On : 28 May 2022 04:49
Operator : JC/MD
Sample : N3022-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
08BR-20220525

Quant Time: May 28 08:39:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

