

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052722\
 Data File : VX029059.D
 Acq On : 28 May 2022 05:59
 Operator : JC/MD
 Sample : PB145106TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB145106TB

Quant Time: May 28 08:39:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	324222	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	577837	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	562739	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	265494	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	214783	48.327	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.660%
35) Dibromofluoromethane	5.391	113	196026	51.778	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	103.560%
50) Toluene-d8	8.653	98	710026	49.744	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.480%
62) 4-Bromofluorobenzene	11.079	95	279275	52.384	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	104.760%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	28178	14.773	ug/l	96
20) Methylene Chloride	2.788	84	21868	5.465	ug/l	96
27) cis-1,2-Dichloroethene	4.495	96	12888	2.920	ug/l	95
43) Isopropyl Acetate	6.342	43	26736	2.877	ug/l	98
44) Trichloroethene	7.129	130	36009	8.124	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052722\
Data File : VX029059.D
Acq On : 28 May 2022 05:59
Operator : JC/MD
Sample : PB145106TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB145106TB

Quant Time: May 28 08:39:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

