


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052722\
 Data File : VX029015.D
 Acq On : 27 May 2022 09:23
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Title : SW846 8260
 Last Update : Thu May 12 14:58:53 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.4	13367	PASS
75	95	30	60	50.6	38856	PASS
95	95	100	100	100.0	76816	PASS
96	95	5	9	6.8	5221	PASS
173	174	0.00	2	1.0	549	PASS
174	95	50	100	74.1	56957	PASS
175	174	5	9	7.7	4395	PASS
176	174	95	101	97.5	55528	PASS
177	176	5	9	6.9	3857	PASS