



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052722\
 Data File : VX029040.D
 Acq On : 27 May 2022 21:50
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Title : SW846 8260
 Last Update : Thu May 12 14:58:53 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.1	12759	PASS
75	95	30	60	48.7	38536	PASS
95	95	100	100	100.0	79200	PASS
96	95	5	9	6.7	5290	PASS
173	174	0.00	2	0.8	457	PASS
174	95	50	100	71.7	56760	PASS
175	174	5	9	7.5	4252	PASS
176	174	95	101	96.7	54912	PASS
177	176	5	9	6.9	3768	PASS