

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052819\
 Data File : VX009920.D
 Acq On : 28 May 2019 17:20
 Operator : JC/SP
 Sample : K3036-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FEQC-FB-052319

Quant Time: May 29 04:57:06 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 29 04:00:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	178893	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	285103	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	256863	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	112258	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	119431	47.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.52%	
35) Dibromofluoromethane	5.50	113	87535	48.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.68%	
50) Toluene-d8	8.71	98	363445	49.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.86%	
62) 4-Bromofluorobenzene	11.13	95	124293	47.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.08%	
Target Compounds						
16) Acetone	2.45	43	9612	6.583	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052819\
Data File : VX009920.D
Acq On : 28 May 2019 17:20
Operator : JC/SP
Sample : K3036-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FEQC-FB-052319

Quant Time: May 29 04:57:06 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
Quant Title : SW846 8260
QLast Update : Wed May 29 04:00:56 2019
Response via : Initial Calibration

