

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052819\
 Data File : VX009921.D
 Acq On : 28 May 2019 17:43
 Operator : JC/SP
 Sample : K3032-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S32-0654

Quant Time: May 29 05:01:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 29 04:00:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	182800	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	295463	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	263386	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	114650	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	123902	47.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.96%	
35) Dibromofluoromethane	5.50	113	90840	48.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.80%	
50) Toluene-d8	8.71	98	375635	49.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.60%	
62) 4-Bromofluorobenzene	11.13	95	126871	46.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.66%	
Target Compounds						
4) Vinyl Chloride	1.41	62	35118	12.972	ug/l	97
16) Acetone	2.45	43	4747	3.182	ug/l	97
27) cis-1,2-Dichloroethene	4.60	96	8771	3.689	ug/l	89
52) Toluene	8.79	92	242567	44.925	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052819\
Data File : VX009921.D
Acq On : 28 May 2019 17:43
Operator : JC/SP
Sample : K3032-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleID :
S32-0654

Quant Time: May 29 05:01:13 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
Quant Title : SW846 8260
QLast Update : Wed May 29 04:00:56 2019
Response via : Initial Calibration

