

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052819\
 Data File : VX009922.D
 Acq On : 28 May 2019 18:07
 Operator : JC/SP
 Sample : K3032-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S32-ER-2019-1

Quant Time: May 29 05:04:00 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 29 04:00:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	180413	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	287960	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	258304	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	113635	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	121315	47.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.20%	
35) Dibromofluoromethane	5.50	113	87344	48.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.50%	
50) Toluene-d8	8.71	98	365652	49.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.48%	
62) 4-Bromofluorobenzene	11.13	95	122793	46.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.00%	
Target Compounds						
20) Methylene Chloride	2.86	84	14145	6.273	ug/l	97
29) Tetrahydrofuran	5.13	42	84226	57.254	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052819\
Data File : VX009922.D
Acq On : 28 May 2019 18:07
Operator : JC/SP
Sample : K3032-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S32-ER-2019-1

Quant Time: May 29 05:04:00 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
Quant Title : SW846 8260
QLast Update : Wed May 29 04:00:56 2019
Response via : Initial Calibration

