

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052819\
 Data File : VX009923.D
 Acq On : 28 May 2019 18:30
 Operator : JC/SP
 Sample : K3015-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OUTFALL001-WK-4

Quant Time: May 29 05:06:49 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 29 04:00:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	182220	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	294028	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	264064	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	111363	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	124460	48.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.70%	
35) Dibromofluoromethane	5.50	113	90626	49.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.06%	
50) Toluene-d8	8.71	98	374664	49.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.82%	
62) 4-Bromofluorobenzene	11.13	95	123862	45.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.88%	
Target Compounds						
16) Acetone	2.45	43	3503	2.355	ug/l	93
27) cis-1,2-Dichloroethene	4.60	96	28509	12.028	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052819\
Data File : VX009923.D
Acq On : 28 May 2019 18:30
Operator : JC/SP
Sample : K3015-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OUTFALL001-WK-4

Quant Time: May 29 05:06:49 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
Quant Title : SW846 8260
QLast Update : Wed May 29 04:00:56 2019
Response via : Initial Calibration

