

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052820\
 Data File : VX016474.D
 Acq On : 28 May 2020 10:59
 Operator : JC/SP
 Sample : L2738-03 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 25-28-COMP

Quant Time: May 29 08:53:50 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 16:31:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.62	168	341682	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	556453	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	544509	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	271963	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	201909	47.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.16%	
35) Dibromofluoromethane	5.46	113	163110	47.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.20%	
50) Toluene-d8	8.70	98	685881	51.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.86%	
62) 4-Bromofluorobenzene	11.12	95	273564	53.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.18%	
Target Compounds						
16) Acetone	2.42	43	51045	27.457	ug/l	98
25) 2-Butanone	4.64	43	39066	12.401	ug/l	100
46) Dibromomethane	7.64	93	3004	1.120	ug/l	97
51) 4-Methyl-2-Pentanone	8.62	43	5116	0.833	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052820\
Data File : VX016474.D
Acq On : 28 May 2020 10:59
Operator : JC/SP
Sample : L2738-03 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
25-28-COMP

Quant Time: May 29 08:53:50 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
Quant Title : SW846 8260
QLast Update : Wed May 27 16:31:05 2020
Response via : Initial Calibration

