

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX052820\
 Data File : VX016481.D
 Acq On : 28 May 2020 13:38
 Operator : JC/SP
 Sample : L2730-05DL 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z2-DUP-0558-200520DL

Manual Integrations
 APPROVED

MMDadoda
 5/29/2020 10:57:49 AM

Quant Time: May 29 06:05:25 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 16:31:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	308376	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	501526	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	493305	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	248930	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	183207	47.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.68%	
35) Dibromofluoromethane	5.47	113	143741	46.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.10%	
50) Toluene-d8	8.70	98	618643	51.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.94%	
62) 4-Bromofluorobenzene	11.12	95	249148	53.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.30%	
Target Compounds						
						Qvalue
11) Tert butyl alcohol	3.01	59	2236	2.277	ug/l #	82
40) Benzene	6.12	78	40282	2.849	ug/l	95
65) Chlorobenzene	10.12	112	610508	60.704	ug/l	100
83) tert-Butylbenzene	11.76	119	8261	0.671	ug/l	90
87) 1,3-Dichlorobenzene	12.01	146	4915	0.627	ug/l	96
88) 1,4-Dichlorobenzene	12.08	146	15856m	1.978	ug/l	
91) 1,2-Dichlorobenzene	12.38	146	25776	3.395	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052820\
Data File : VX016481.D
Acq On : 28 May 2020 13:38
Operator : JC/SP
Sample : L2730-05DL 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z2-DUP-0558-200520DL

Manual Integrations
APPROVED

MMDadoda
5/29/2020 10:57:49 AM

Quant Time: May 29 06:05:25 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
Quant Title : SW846 8260
QLast Update : Wed May 27 16:31:05 2020
Response via : Initial Calibration

