

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052820\
 Data File : VX016482.D
 Acq On : 28 May 2020 14:01
 Operator : JC/SP
 Sample : L2786-01 20X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LD-C

Quant Time: May 28 14:44:14 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 16:31:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	289481	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	474077	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	482884	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	252312	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.03	65	176115	48.99	ug/l	0.00
Spiked Amount			Recovery	=	97.98%	
35) Dibromofluoromethane	5.46	113	144822	49.61	ug/l	0.00
Spiked Amount			Recovery	=	99.22%	
50) Toluene-d8	8.70	98	591262	52.04	ug/l	0.00
Spiked Amount			Recovery	=	104.08%	
62) 4-Bromofluorobenzene	11.12	95	249856	56.91	ug/l	0.00
Spiked Amount			Recovery	=	113.82%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.43	43	53800	34.158	ug/l	96
17) Carbon Disulfide	2.54	76	3625191	438.264	ug/l	100
40) Benzene	6.12	78	13394	1.002	ug/l #	85
52) Toluene	8.77	92	2718	0.322	ug/l	90
65) Chlorobenzene	10.12	112	9545	0.970	ug/l	97
68) m/p-Xylenes	10.34	106	3565	0.549	ug/l	96
69) o-Xylene	10.68	106	1561	0.248	ug/l	80
95) Naphthalene	13.82	128	126472	7.218	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052820\
Data File : VX016482.D
Acq On : 28 May 2020 14:01
Operator : JC/SP
Sample : L2786-01 20X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LD-C

Quant Time: May 28 14:44:14 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
Quant Title : SW846 8260
QLast Update : Wed May 27 16:31:05 2020
Response via : Initial Calibration

