

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052820\
 Data File : VX016488.D
 Acq On : 28 May 2020 16:17
 Operator : JC/SP
 Sample : L2782-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-01-200526

Quant Time: May 29 09:06:45 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 16:31:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	284933	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	455733	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	456959	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	228032	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	171509	48.47	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.94%	
35) Dibromofluoromethane	5.46	113	133241	47.48	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.96%	
50) Toluene-d8	8.70	98	578327	52.95	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.90%	
62) 4-Bromofluorobenzene	11.12	95	231640	54.89	ug/l	0.00
Spiked Amount	50.000		Recovery	=	109.78%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052820\
Data File : VX016488.D
Acq On : 28 May 2020 16:17
Operator : JC/SP
Sample : L2782-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-01-200526

Quant Time: May 29 09:06:45 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
Quant Title : SW846 8260
QLast Update : Wed May 27 16:31:05 2020
Response via : Initial Calibration

