

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052820\
 Data File : VX016491.D
 Acq On : 28 May 2020 17:25
 Operator : JC/SP
 Sample : L2767-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRE-1006

Quant Time: May 29 05:50:20 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 16:31:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	270725	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	443418	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	440168	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	230063	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	164683	48.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.96%	
35) Dibromofluoromethane	5.47	113	129253	47.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.68%	
50) Toluene-d8	8.70	98	548202	51.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.18%	
62) 4-Bromofluorobenzene	11.12	95	226291	55.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.22%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	3.01	59	6611	7.668	ug/l	99
16) Acetone	2.42	43	79622	54.055	ug/l	98
20) Methylene Chloride	2.83	84	9573	3.141	ug/l	94
25) 2-Butanone	4.64	43	51215	20.519	ug/l	99
51) 4-Methyl-2-Pentanone	8.62	43	49812	10.175	ug/l	99
80) 1,3,5-Trimethylbenzene	11.49	105	17992	1.352	ug/l	97
84) 1,2,4-Trimethylbenzene	11.79	105	64360	4.781	ug/l	100
86) p-Isopropyltoluene	12.05	119	46291	3.464	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052820\
Data File : VX016491.D
Acq On : 28 May 2020 17:25
Operator : JC/SP
Sample : L2767-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRE-1006

Quant Time: May 29 05:50:20 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
Quant Title : SW846 8260
QLast Update : Wed May 27 16:31:05 2020
Response via : Initial Calibration

