

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052820\
 Data File : VX016492.D
 Acq On : 28 May 2020 17:48
 Operator : JC/SP
 Sample : L2777-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS036MW002-200526

Quant Time: May 29 05:41:32 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 16:31:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	270613	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	445840	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	445460	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	220054	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	160043	47.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.24%	
35) Dibromofluoromethane	5.47	113	126844	46.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.42%	
50) Toluene-d8	8.70	98	549891	51.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.92%	
62) 4-Bromofluorobenzene	11.12	95	227315	55.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.12%	
Target Compounds						
16) Acetone	2.42	43	6301	4.279	ug/l	96
17) Carbon Disulfide	2.56	76	2800	0.362	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052820\
Data File : VX016492.D
Acq On : 28 May 2020 17:48
Operator : JC/SP
Sample : L2777-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS036MW002-200526

Quant Time: May 29 05:41:32 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
Quant Title : SW846 8260
QLast Update : Wed May 27 16:31:05 2020
Response via : Initial Calibration

