

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052820\
 Data File : VX016494.D
 Acq On : 28 May 2020 18:34
 Operator : JC/SP
 Sample : L2777-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY070MW233-200526

Quant Time: May 29 04:06:45 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 16:31:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	278361	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	454425	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	458541	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	226901	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	166417	48.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.28%	
35) Dibromofluoromethane	5.47	113	130909	46.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.56%	
50) Toluene-d8	8.70	98	569213	52.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.54%	
62) 4-Bromofluorobenzene	11.12	95	232892	55.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.68%	
Target Compounds						
4) Vinyl Chloride	1.39	62	4068	1.162	ug/l	94
16) Acetone	2.42	43	38987	25.742	ug/l	99
17) Carbon Disulfide	2.55	76	2835	0.356	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052820\
Data File : VX016494.D
Acq On : 28 May 2020 18:34
Operator : JC/SP
Sample : L2777-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY070MW233-200526

Quant Time: May 29 04:06:45 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
Quant Title : SW846 8260
QLast Update : Wed May 27 16:31:05 2020
Response via : Initial Calibration

