

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052821\
 Data File : VX022225.D
 Acq On : 27 May 2021 12:22
 Operator : JC/MD
 Sample : VX0528MBL01
 Misc : 5.0g/10mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0528MBL01

Quant Time: May 27 13:40:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	67830	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	146313	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	127515	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	49103	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	57631	55.08	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	110.16%
35) Dibromofluoromethane	5.391	113	44947	48.44	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.88%
50) Toluene-d8	8.653	98	176438	49.36	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.72%
62) 4-Bromofluorobenzene	11.085	95	61153	46.29	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.58%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052821\
Data File : VX022225.D
Acq On : 27 May 2021 12:22
Operator : JC/MD
Sample : VX0528MBL01
Misc : 5.0g/10mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0528MBL01

Quant Time: May 27 13:40:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

