

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052821\
 Data File : VX022228.D
 Acq On : 27 May 2021 13:32
 Operator : JC/MD
 Sample : M2535-01 20X
 Misc : 6.47g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GHOST-1D

Manual Integrations
 APPROVED

MMDadoda
 5/28/2021 2:37:41 PM

Quant Time: May 27 14:07:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	66826	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	137670	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	124874	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	56203	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	52655	51.08	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 102.16%			
35) Dibromofluoromethane	5.397	113	41416	47.44	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 94.88%			
50) Toluene-d8	8.653	98	175896	52.30	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 104.60%			
62) 4-Bromofluorobenzene	11.085	95	64158	51.61	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 103.22%			
Target Compounds						Qvalue
31) Cyclohexane	5.483	56	83532	64.48	ug/l	# 87
39) Methylcyclohexane	7.385	83	230926	159.94	ug/l	96
40) Benzene	6.050	78	72080	18.40	ug/l	95
52) Toluene	8.726	92	1281754	516.73	ug/l	98
67) Ethyl Benzene	10.201	91	2635379	551.10	ug/l	99
68) m/p-Xylenes	10.311	106	3558639	1969.06	ug/l	95
69) o-Xylene	10.647	106	1571335	893.96	ug/l	99
73) Isopropylbenzene	10.964	105	366164	76.98	ug/l	100
78) n-propylbenzene	11.311	91	1620005	290.76	ug/l	98
80) 1,3,5-Trimethylbenzene	11.457	105	1829603	464.07	ug/l	100
84) 1,2,4-Trimethylbenzene	11.762	105	5780110	1456.55	ug/l	97
86) p-Isopropyltoluene	12.012	119	74183m	18.28	ug/l	
89) n-Butylbenzene	12.335	91	596058	158.78	ug/l	# 45
95) Naphthalene	13.780	128	1219233	277.38	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052821\
Data File : VX022228.D
Acq On : 27 May 2021 13:32
Operator : JC/MD
Sample : M2535-01 20X
Misc : 6.47g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GHOST-1D

Manual Integrations
APPROVED

MMDadoda
5/28/2021 2:37:41 PM

Quant Time: May 27 14:07:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

