

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052821\
 Data File : VX022230.D
 Acq On : 27 May 2021 14:19
 Operator : JC/MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VIBLK

Quant Time: May 27 14:45:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	69574	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	144464	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	126157	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	49519	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	56740	52.87	ug/l	0.00
Spiked Amount 50.000	Range	78 - 117	Recovery	= 105.74%		
35) Dibromofluoromethane	5.403	113	43085	47.03	ug/l	0.01
Spiked Amount 50.000	Range	75 - 124	Recovery	= 94.06%		
50) Toluene-d8	8.653	98	176922	50.13	ug/l	0.00
Spiked Amount 50.000	Range	92 - 112	Recovery	= 100.26%		
62) 4-Bromofluorobenzene	11.085	95	60794	46.61	ug/l	0.00
Spiked Amount 50.000	Range	83 - 123	Recovery	= 93.22%		
Target Compounds						
						Qvalue
68) m/p-Xylenes	10.305	106	1751	0.96	ug/l	92
84) 1,2,4-Trimethylbenzene	11.756	105	2947	0.84	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052821\
Data File : VX022230.D
Acq On : 27 May 2021 14:19
Operator : JC/MD
Sample : VIBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VIBLK

Quant Time: May 27 14:45:54 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

