

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052821\
 Data File : VX022233.D
 Acq On : 27 May 2021 15:30
 Operator : JC/MD
 Sample : M2535-01DL 800X
 Misc : 6.47g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GHOST-1DDL

Manual Integrations
 APPROVED

MMDadoda
 5/28/2021 2:37:48 PM

Quant Time: May 27 16:04:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	67558	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	138609	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	120073	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	44149	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	54724	52.51	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 105.02%			
35) Dibromofluoromethane	5.397	113	42131	47.93	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 95.86%			
50) Toluene-d8	8.653	98	170804	50.44	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 100.88%			
62) 4-Bromofluorobenzene	11.085	95	57010	45.55	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 91.10%			
Target Compounds					Qvalue	
31) Cyclohexane	5.489	56	1801m	1.38	ug/l	
39) Methylcyclohexane	7.385	83	5171	3.56	ug/l #	80
40) Benzene	6.056	78	1707	0.43	ug/l #	90
52) Toluene	8.720	92	27846	11.15	ug/l	100
67) Ethyl Benzene	10.195	91	56494	12.29	ug/l	97
68) m/p-Xylenes	10.305	106	84241	48.48	ug/l	98
69) o-Xylene	10.646	106	34726	20.55	ug/l	98
73) Isopropylbenzene	10.963	105	7195	1.93	ug/l	98
78) n-propylbenzene	11.305	91	31246	7.14	ug/l	99
80) 1,3,5-Trimethylbenzene	11.457	105	37990	12.27	ug/l	99
84) 1,2,4-Trimethylbenzene	11.756	105	139913	44.88	ug/l	99
86) p-Isopropyltoluene	12.012	119	1856	0.58	ug/l	89
89) n-Butylbenzene	12.335	91	10902	3.70	ug/l #	44
95) Naphthalene	13.780	128	20923	6.06	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052821\
Data File : VX022233.D
Acq On : 27 May 2021 15:30
Operator : JC/MD
Sample : M2535-01DL 800X
Misc : 6.47g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GHOST-1DDL

Manual Integrations
APPROVED

MMDadoda
5/28/2021 2:37:48 PM

Quant Time: May 27 16:04:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

