

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052821\
 Data File : VX022238.D
 Acq On : 27 May 2021 17:27
 Operator : JC/MD
 Sample : M2535-05
 Misc : 5.99g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GHOST-5D

Quant Time: May 27 17:56:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	70346	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	145379	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	134848	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.030	152	68042	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	56291	51.87	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.74%
35) Dibromofluoromethane	5.391	113	42286	45.87	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.74%
50) Toluene-d8	8.653	98	183667	51.72	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	103.44%
62) 4-Bromofluorobenzene	11.085	95	72858	55.50	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	111.00%
Target Compounds						
					Qvalue	
31) Cyclohexane	5.464	56	30297	22.22	ug/l	89
39) Methylcyclohexane	7.379	83	113134	74.20	ug/l	97
40) Benzene	6.031	78	11636	2.81	ug/l #	87
67) Ethyl Benzene	10.195	91	115028	22.28	ug/l	100
68) m/p-Xylenes	10.305	106	19586	10.04	ug/l	100
73) Isopropylbenzene	10.969	105	135127	23.46	ug/l	98
78) n-propylbenzene	11.311	91	666931	98.87	ug/l	97
80) 1,3,5-Trimethylbenzene	11.457	105	82558	17.30	ug/l	99
84) 1,2,4-Trimethylbenzene	11.756	105	1095261	227.98	ug/l	99
85) sec-Butylbenzene	11.896	105	107780	18.53	ug/l	92
86) p-Isopropyltoluene	12.012	119	37294	7.59	ug/l	92
89) n-Butylbenzene	12.335	91	355251	78.17	ug/l #	65
95) Naphthalene	13.780	128	470237	88.37	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052821\
Data File : VX022238.D
Acq On : 27 May 2021 17:27
Operator : JC/MD
Sample : M2535-05
Misc : 5.99g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GHOST-5D

Quant Time: May 27 17:56:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

