

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052821\
 Data File : VX022240.D
 Acq On : 27 May 2021 18:14
 Operator : JC/MD
 Sample : M2526-18DL 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 INTERCEPTOR-SUMP-20210525DL

Quant Time: May 28 01:33:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	68534	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	142074	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	124250	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	46299	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	55535	52.53	ug/l	0.00
Spiked Amount 50.000	Range	78 - 117	Recovery	=	105.06%	
35) Dibromofluoromethane	5.391	113	42323	46.98	ug/l	0.00
Spiked Amount 50.000	Range	75 - 124	Recovery	=	93.96%	
50) Toluene-d8	8.653	98	174590	50.31	ug/l	0.00
Spiked Amount 50.000	Range	92 - 112	Recovery	=	100.62%	
62) 4-Bromofluorobenzene	11.085	95	58626	45.70	ug/l	0.00
Spiked Amount 50.000	Range	83 - 123	Recovery	=	91.40%	
Target Compounds						
						Qvalue
12) 1,1-Dichloroethene	2.325	96	362	0.46	ug/l	# 48
21) trans-1,2-Dichloroethene	3.105	96	466	0.55	ug/l	86
27) cis-1,2-Dichloroethene	4.495	96	14751	14.20	ug/l	90
44) Trichloroethene	7.129	130	60150	59.88	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052821\
Data File : VX022240.D
Acq On : 27 May 2021 18:14
Operator : JC/MD
Sample : M2526-18DL 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
INTERCEPTOR-SUMP-20210525DL

Quant Time: May 28 01:33:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

