

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052919\
 Data File : VX009932.D
 Acq On : 29 May 2019 12:09
 Operator : JC/SP
 Sample : IBLK
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: May 30 01:44:24 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 29 04:00:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	184984	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	296969	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	264851	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	115565	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	124698	47.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.44%	
35) Dibromofluoromethane	5.50	113	90483	48.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.92%	
50) Toluene-d8	8.71	98	374629	49.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.82%	
62) 4-Bromofluorobenzene	11.13	95	125708	46.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.32%	
Target Compounds						
4) Vinyl Chloride	1.41	62	4266	1.557	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052919\
Data File : VX009932.D
Acq On : 29 May 2019 12:09
Operator : JC/SP
Sample : IBLK
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: May 30 01:44:24 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
Quant Title : SW846 8260
QLast Update : Wed May 29 04:00:56 2019
Response via : Initial Calibration

