

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052919\
 Data File : VX009935.D
 Acq On : 29 May 2019 13:19
 Operator : JC/SP
 Sample : K3038-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LF012MW001-190523

Quant Time: May 30 01:52:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 29 04:00:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	181697	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	290838	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	257000	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	111702	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	122803	47.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.70%	
35) Dibromofluoromethane	5.50	113	90151	49.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.60%	
50) Toluene-d8	8.71	98	364925	49.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.30%	
62) 4-Bromofluorobenzene	11.13	95	123406	46.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.54%	
Target Compounds						
4) Vinyl Chloride	1.41	62	16631	6.181	ug/l	100
27) cis-1,2-Dichloroethene	4.60	96	249705	105.655	ug/l	88
65) Chlorobenzene	10.13	112	4264	0.793	ug/l	95
91) 1,2-Dichlorobenzene	12.39	146	2026	0.550	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052919\
Data File : VX009935.D
Acq On : 29 May 2019 13:19
Operator : JC/SP
Sample : K3038-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LF012MW001-190523

Quant Time: May 30 01:52:57 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
Quant Title : SW846 8260
QLast Update : Wed May 29 04:00:56 2019
Response via : Initial Calibration

