

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052919\
 Data File : VX009943.D
 Acq On : 29 May 2019 16:26
 Operator : JC/SP
 Sample : K3088-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK

Quant Time: May 30 02:21:35 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 29 04:00:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	173569	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	280733	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	248833	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	108703	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	118515	48.34	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.68%	
35) Dibromofluoromethane	5.50	113	86502	49.01	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.02%	
50) Toluene-d8	8.71	98	355562	49.61	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.22%	
62) 4-Bromofluorobenzene	11.13	95	119658	46.48	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.96%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052919\
Data File : VX009943.D
Acq On : 29 May 2019 16:26
Operator : JC/SP
Sample : K3088-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK

Quant Time: May 30 02:21:35 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
Quant Title : SW846 8260
QLast Update : Wed May 29 04:00:56 2019
Response via : Initial Calibration

