

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX052920\
 Data File : VX016508.D
 Acq On : 29 May 2020 14:09
 Operator : JC/SP
 Sample : L2778-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WP022RW349-200526

Quant Time: May 30 01:42:17 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 16:31:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	284622	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	459861	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	448443	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	218033	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.03	65	169563	47.97	ug/l	0.00
Spiked Amount			Recovery	=	95.94%	
35) Dibromofluoromethane	5.47	113	133505	47.15	ug/l	0.00
Spiked Amount			Recovery	=	94.30%	
50) Toluene-d8	8.70	98	566424	51.39	ug/l	0.00
Spiked Amount			Recovery	=	102.78%	
62) 4-Bromofluorobenzene	11.12	95	225008	52.84	ug/l	0.00
Spiked Amount			Recovery	=	105.68%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) Chloromethane	1.31	50	2336	0.690	ug/l	99
4) Vinyl Chloride	1.39	62	288558	80.612	ug/l	99
10) Methyl Iodide	2.50	142	274	3.139	ug/l #	90
11) Tert butyl alcohol	3.00	59	1619	1.786	ug/l #	64
12) 1,1-Dichloroethene	2.36	96	4056	1.481	ug/l	95
13) Acrolein	2.26	56	379	0.983	ug/l	99
16) Acetone	2.42	43	10070	6.503	ug/l	97
17) Carbon Disulfide	2.56	76	1920	0.236	ug/l #	91
18) Methyl Acetate	2.74	43	918	0.231	ug/l #	75
21) trans-1,2-Dichloroethene	3.15	96	4826	1.593	ug/l	87
24) 1,1-Dichloroethane	3.68	63	12839	2.351	ug/l	96
25) 2-Butanone	4.64	43	1101	0.420	ug/l #	84
27) cis-1,2-Dichloroethene	4.57	96	258746	73.833	ug/l	92
36) 1,1-Dichloropropene	5.63	75	20504	4.709	ug/l #	51
38) Carbon Tetrachloride	5.63	117	26486	5.789	ug/l #	16
40) Benzene	6.12	78	9887	0.763	ug/l	96
49) 1,4-Dioxane	7.73	88	25	0.274	ug/l #	1
52) Toluene	8.77	92	3079	0.377	ug/l	100
65) Chlorobenzene	10.12	112	136143	14.891	ug/l	100
76) 1,2,3-Trichloropropane	11.12	75	110980	23.157	ug/l #	35
81) trans-1,4-Dichloro-2-buten	11.12	75	110980	56.963	ug/l #	14
87) 1,3-Dichlorobenzene	12.01	146	12737	1.855	ug/l	99
88) 1,4-Dichlorobenzene	12.08	146	44748	6.374	ug/l	96
91) 1,2-Dichlorobenzene	12.38	146	147453	22.171	ug/l	99
93) 1,2,4-Trichlorobenzene	13.63	180	1677	0.380	ug/l	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052920\
Data File : VX016508.D
Acq On : 29 May 2020 14:09
Operator : JC/SP
Sample : L2778-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WP022RW349-200526

Quant Time: May 30 01:42:17 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
Quant Title : SW846 8260
QLast Update : Wed May 27 16:31:05 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.64 min Scan# 746

Delta R.T. 0.01 min

Lab File: VX016508.D

Acq: 29 May 2020 14:09

Instrument :

MSVOA_X

ClientSampleId :


WP022RW349-200526

Tot Ion:168 Resp: 284622

Ion Ratio Lower Upper




168 100

99 52.6 45.2 67.8



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#3

Chloromethane

Concen: 0.690 ug/l

RT: 1.31 min Scan# 37

Delta R.T. -0.00 min

Lab File: VX016508.D


Acq: 29 May 2020 14:09

Tgt Ion: 50 Resp: 2336

Ion Ratio Lower Upper



50 100


52 32.3 26.2 39.2



Abundance Ion 49.90 (49.60 to 50.60): VX0

Ion 51.90 (51.60 to 52.60): VX0





#4

Vinyl Chloride

Concen: 80.612 ug/l

RT: 1.39 min Scan# 50

Delta R.T. -0.00 min

Lab File: VX016508.D

Acq: 29 May 2020 14:09

Instrument :

MSVOA_X

ClientSampleId :


WP022RW349-200526

Tot Ion: 62 Resp: 288558

Ion Ratio Lower Upper




62 100

64 31.6 25.9 38.9



Abundance Ion 61.90 (61.60 to 62.60): VX0

Ion 63.90 (63.60 to 64.60): VX0



#10

Methyl Iodide

Concen: 3.139 ug/l

RT: 2.50 min Scan# 232

Delta R.T. 0.01 min

Lab File: VX016508.D

Acq: 29 May 2020 14:09


Tgt Ion:142 Resp: 274

Ion Ratio Lower Upper

142 100

127 50.7 35.3 52.9



141 17.5 11.4 17.2#




Abundance Ion 141.80 (141.50 to 142.50): V

Ion 126.80 (126.50 to 127.50): V

Ion 140.80 (140.50 to 141.50): V






#11




Tert butyl alcohol
 Concen: 1.786 ug/l
 RT: 3.00 min Scan# 314
 Delta R.T. -0.02 min
 Lab File: VX016508.D
 Acq: 29 May 2020 14:09

Instrument :
 MSVOA_X
 ClientSampleId :
 WP022RW349-200526

Tot Ion: 59 Resp: 1619
 Ion Ratio Lower Upper
 59 100
 57 23.7 8.2 12.4#




Abundance Ion 59.00 (58.70 to 59.70): VX0
 Ion 57.00 (56.70 to 57.70): VX0





#12


1,1-Dichloroethene
 Concen: 1.481 ug/l
 RT: 2.36 min Scan# 209
 Delta R.T. 0.01 min
 Lab File: VX016508.D
 Acq: 29 May 2020 14:09

Tgt Ion: 96 Resp: 4056
 Ion Ratio Lower Upper
 96 100
 61 151.5 126.6 189.8
 98 67.5 51.3 76.9



Abundance Ion 95.90 (95.60 to 96.60): VX0
 Ion 60.90 (60.60 to 61.60): VX0
 Ion 97.90 (97.60 to 98.60): VX0





#13

Acrolein

Concen: 0.983 ug/l

RT: 2.26 min Scan# 193

Delta R.T. -0.01 min

Lab File: VX016508.D

Acq: 29 May 2020 14:09

Instrument :

MSVOA_X

ClientSampleId :





WP022RW349-200526

Tot Ion: 56 Resp: 379

Ion Ratio Lower Upper

56 100

55 71.8 58.3 87.5



#16

Acetone

Concen: 6.503 ug/l

RT: 2.42 min Scan# 218

Delta R.T. -0.00 min

Lab File: VX016508.D




Acq: 29 May 2020 14:09


Tgt Ion: 43 Resp: 10070

Ion Ratio Lower Upper

43 100

58 30.8 26.1 39.1





#17

Carbon Disulfide

Concen: 0.236 ug/l

RT: 2.56 min Scan# 241

Delta R.T. 0.01 min

Lab File: VX016508.D

Acq: 29 May 2020 14:09

Instrument :

MSVOA_X

ClientSampleId :





WP022RW349-200526

Tot Ion: 76 Resp: 1920

Ion Ratio Lower Upper

76 100

78 5.8 7.4 11.0#



#18

Methyl Acetate

Concen: 0.231 ug/l

RT: 2.74 min Scan# 271

Delta R.T. -0.01 min

Lab File: VX016508.D




Acq: 29 May 2020 14:09


Tgt Ion: 43 Resp: 918

Ion Ratio Lower Upper

43 100

74 12.2 19.7 29.5#





#21

trans-1,2-Dichloroethene

Concen: 1.593 ug/l

RT: 3.15 min Scan# 338

Delta R.T. 0.01 min

Lab File: VX016508.D

Acq: 29 May 2020 14:09

Instrument :

MSVOA_X

ClientSampleId :

WP022RW349-200526





Tot Ion: 96 Resp: 4826

Ion Ratio Lower Upper

96 100

61 128.5 115.3 172.9

98 54.9 51.9 77.9



#24

1,1-Dichloroethane

Concen: 2.351 ug/l

RT: 3.68 min Scan# 425

Delta R.T. 0.01 min

Lab File: VX016508.D

Acq: 29 May 2020 14:09




Tgt Ion: 63 Resp: 12839


Ion Ratio Lower Upper

63 100

98 8.1 3.6 10.9

100 7.0 2.4 7.1





#25

2-Butanone

Concen: 0.420 ug/l

RT: 4.64 min Scan# 583

Delta R.T. 0.01 min

Lab File: VX016508.D

Acq: 29 May 2020 14:09

Instrument :

MSVOA_X

ClientSampleId :





WP022RW349-200526

Tot Ion: 43 Resp: 1101

Ion Ratio Lower Upper

43 100

72 34.4 21.0 31.6#



#27

cis-1,2-Dichloroethene

Concen: 73.833 ug/l

RT: 4.57 min Scan# 571

Delta R.T. 0.01 min

Lab File: VX016508.D

Acq: 29 May 2020 14:09




Tgt Ion: 96 Resp: 258746


Ion Ratio Lower Upper

96 100

61 128.2 0.0 284.6

98 64.6 0.0 127.2





#33

1,2-Dichloroethane-d4

Concen: 47.972 ug/l

RT: 6.03 min Scan# 811

Delta R.T. -0.00 min

Lab File: VX016508.D

Acq: 29 May 2020 14:09

Instrument :

MSVOA_X

ClientSampleId :





WP022RW349-200526

Tot Ion: 65 Resp: 169563

Ion Ratio Lower Upper

65 100

67 52.9 0.0 104.8



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.84 min Scan# 943

Delta R.T. 0.01 min

Lab File: VX016508.D

Acq: 29 May 2020 14:09




Tgt Ion:114 Resp: 459861


Ion Ratio Lower Upper

114 100

63 20.1 0.0 42.6

88 15.8 0.0 32.8









#35
Dibromofluoromethane
Concen: 47.149 ug/l
RT: 5.47 min Scan# 719
Delta R.T. 0.01 min
Lab File: VX016508.D
Acq: 29 May 2020 14:09




Instrument :
MSVOA_X
ClientSampleId :
WP022RW349-200526


Tot Ion:113 Resp: 133505
Ion Ratio Lower Upper
113 100
111 102.3 83.0 124.6
192 22.3 17.7 26.5



#36
1,1-Dichloropropene
Concen: 4.709 ug/l
RT: 5.63 min Scan# 745
Delta R.T. -0.14 min
Lab File: VX016508.D
Acq: 29 May 2020 14:09

Tgt Ion: 75 Resp: 20504
Ion Ratio Lower Upper
75 100
110 9.4 17.4 52.3#
77 0.1 24.2 36.4#





#38

Carbon Tetrachloride

Concen: 5.789 ug/l

RT: 5.63 min Scan# 745

Delta R.T. -0.13 min

Lab File: VX016508.D

Acq: 29 May 2020 14:09

Instrument :

MSVOA_X

ClientSampleId :

WP022RW349-200526


Tot Ion:117 Resp: 26486

Ion Ratio Lower Upper

117 100

119 5.5 76.3 114.5#




121 0.0 24.6 37.0#



Abundance Ion 116.80 (116.50 to 117.50): V

Ion 118.80 (118.50 to 119.50): V

Ion 120.80 (120.50 to 121.50): V



#40

Benzene

Concen: 0.763 ug/l

RT: 6.12 min Scan# 826

Delta R.T. 0.01 min

Lab File: VX016508.D


Acq: 29 May 2020 14:09

Tgt Ion: 78 Resp: 9887

Ion Ratio Lower Upper



78 100


77 25.7 18.9 28.3



Abundance Ion 78.00 (77.70 to 78.70): VX0

Ion 77.00 (76.70 to 77.70): VX0





#49

1,4-Dioxane

Concen: 0.274 ug/l

RT: 7.73 min Scan# 1090

Delta R.T. 0.02 min

Lab File: VX016508.D

Acq: 29 May 2020 14:09

Instrument :

MSVOA_X

ClientSampleId :

WP022RW349-200526


Tot Ion: 88 Resp: 25

Ion Ratio Lower Upper

88 100

43 1288.0 24.2 36.4#


58 0.0 56.5 84.7#





Abundance Ion 88.00 (87.70 to 88.70): VX0

Ion 43.00 (42.70 to 43.70): VX0

Ion 58.00 (57.70 to 58.70): VX0



Time-->



#50

Toluene-d8

Concen: 51.395 ug/l

RT: 8.70 min Scan# 1248

Delta R.T. -0.00 min

Lab File: VX016508.D


Acq: 29 May 2020 14:09

Tgt Ion: 98 Resp: 566424

Ion Ratio Lower Upper


98 100

100 66.9 53.8 80.6





Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): VX0



Time-->





#52

Toluene

Concen: 0.377 ug/l

RT: 8.77 min Scan# 1260

Delta R.T. -0.00 min

Lab File: VX016508.D

Acq: 29 May 2020 14:09

Instrument :

MSVOA_X

ClientSampleId :


WP022RW349-200526

Tot Ion: 92 Resp: 3079

Ion Ratio Lower Upper

92 100

91 169.2 135.0 202.4



Abundance Ion 92.00 (91.70 to 92.70): VX0

Ion 91.00 (90.70 to 91.70): VX0

3000



2000

1000

0

8.77

Time-->



#62

4-Bromofluorobenzene

Concen: 52.837 ug/l

RT: 11.12 min Scan# 1646

Delta R.T. -0.00 min

Lab File: VX016508.D

Acq: 29 May 2020 14:09


Tgt Ion: 95 Resp: 225008

Ion Ratio Lower Upper

95 100

174 81.2 0.0 163.0

176 79.5 0.0 156.8



Abundance Ion 94.90 (94.60 to 95.60): VX0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V

250000

200000

150000


100000


50000

0

11.12

Time-->









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.10 min Scan# 1478
Delta R.T. -0.00 min
Lab File: VX016508.D
Acq: 29 May 2020 14:09




Instrument :
MSVOA_X
ClientSampleId :
WP022RW349-200526


Tot Ion:117 Resp: 448443
Ion Ratio Lower Upper
117 100
82 56.6 45.0 67.6
119 31.4 25.8 38.8



#65
Chlorobenzene
Concen: 14.891 ug/l
RT: 10.12 min Scan# 1482
Delta R.T. -0.00 min
Lab File: VX016508.D
Acq: 29 May 2020 14:09

Tgt Ion:112 Resp: 136143
Ion Ratio Lower Upper
112 100
114 32.4 26.0 39.0





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.06 min Scan# 1800

Delta R.T. -0.00 min

Lab File: VX016508.D

Acq: 29 May 2020 14:09

Instrument :

MSVOA_X

ClientSampleId :

WP022RW349-200526


Tot Ion:152 Resp: 218033

Ion Ratio Lower Upper

152 100

115 58.8 39.9 119.6




150 160.1 0.0 341.0



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V



#76

1,2,3-Trichloropropane

Concen: 23.157 ug/l

RT: 11.12 min Scan# 1646

Delta R.T. -0.16 min

Lab File: VX016508.D


Acq: 29 May 2020 14:09

Tgt Ion: 75 Resp: 110980

Ion Ratio Lower Upper



75 100


77 1.2 21.3 63.7#



Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 77.00 (76.70 to 77.70): VX0





#81

trans-1,4-Dichloro-2-butene

Concen: 56.963 ug/l

RT: 11.12 min Scan# 1646

Delta R.T. 0.06 min

Lab File: VX016508.D

Acq: 29 May 2020 14:09

Instrument :

MSVOA_X

ClientSampleId :

WP022RW349-200526


Tot Ion: 75 Resp: 110980

Ion Ratio Lower Upper

75 100

53 0.1 71.3 106.9#




89 0.0 36.6 55.0#



Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 53.00 (52.70 to 53.70): VX0

Ion 88.90 (88.60 to 89.60): VX0



#87

1,3-Dichlorobenzene

Concen: 1.855 ug/l

RT: 12.01 min Scan# 1792

Delta R.T. -0.00 min

Lab File: VX016508.D

Acq: 29 May 2020 14:09


Tgt Ion: 146 Resp: 12737

Ion Ratio Lower Upper

146 100

111 40.5 20.0 60.0



148 64.5 31.9 95.5




Abundance Ion 145.90 (145.60 to 146.60): V

Ion 110.90 (110.60 to 111.60): V

Ion 147.90 (147.60 to 148.60): V





#88

1,4-Dichlorobenzene

Concen: 6.374 ug/l

RT: 12.08 min Scan# 1803

Delta R.T. -0.00 min

Lab File: VX016508.D

Acq: 29 May 2020 14:09

Instrument :

MSVOA_X

ClientSampleId :

WP022RW349-200526


Tot Ion:146 Resp: 44748

Ion Ratio Lower Upper




146 100

111 44.2 19.4 58.2

148 64.0 31.4 94.3



Abundance Ion 145.90 (145.60 to 146.60): V
Ion 110.90 (110.60 to 111.60): V
Ion 147.90 (147.60 to 148.60): V



#91

1,2-Dichlorobenzene

Concen: 22.171 ug/l

RT: 12.38 min Scan# 1852

Delta R.T. -0.00 min

Lab File: VX016508.D

Acq: 29 May 2020 14:09


Tgt Ion:146 Resp: 147453

Ion Ratio Lower Upper



146 100


111 42.4 21.4 64.3

148 65.1 31.9 95.7



Abundance Ion 145.90 (145.60 to 146.60): V
Ion 110.90 (110.60 to 111.60): V
Ion 147.90 (147.60 to 148.60): V





#93

1,2,4-Trichlorobenzene

Concen: 0.380 ug/l

RT: 13.63 min Scan# 2057

Delta R.T. -0.00 min

Lab File: VX016508.D

Acq: 29 May 2020 14:09

Instrument :

MSVOA_X

ClientSampleId :

WP022RW349-200526


Tot Ion:180 Resp: 1677

Ion Ratio Lower Upper

180 100

182 99.3 48.3 144.8

145 38.0 15.4 46.2



Abundance Ion 179.80 (179.50 to 180.50): V

Ion 181.80 (181.50 to 182.50): V

Ion 144.80 (144.50 to 145.50): V

