

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX052920\
 Data File : VX016512.D
 Acq On : 29 May 2020 15:40
 Operator : JC/SP
 Sample : L2503-15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OR-03-052820

Quant Time: May 30 06:30:46 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 16:31:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	292627	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	471703	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	469232	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	231130	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	170381	46.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.76%	
35) Dibromofluoromethane	5.47	113	138144	47.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.12%	
50) Toluene-d8	8.70	98	586694	51.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.80%	
62) 4-Bromofluorobenzene	11.12	95	237227	54.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.62%	
Target Compounds						
16) Acetone	2.42	43	25287	15.88	ug/l	95
20) Methylene Chloride	2.84	84	4669	1.42	ug/l	92
25) 2-Butanone	4.64	43	11408	4.23	ug/l #	87
43) Isopropyl Acetate	6.42	43	66444	7.94	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX052920\
Data File : VX016512.D
Acq On : 29 May 2020 15:40
Operator : JC/SP
Sample : L2503-15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OR-03-052820

Quant Time: May 30 06:30:46 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
Quant Title : SW846 8260
QLast Update : Wed May 27 16:31:05 2020
Response via : Initial Calibration

