


Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX052920\
 Data File : VX016500.D
 Acq On : 29 May 2020 09:17
 Operator : JC/SP
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Title : SW846 8260
 Last Update : Wed May 27 16:31:05 2020



AutoFind: Scans 1645, 1646, 1647; Background Corrected with Scan 1638

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.0	12891	PASS
75	95	30	60	49.1	37184	PASS
95	95	100	100	100.0	75728	PASS
96	95	5	9	6.3	4773	PASS
173	174	0.00	2	0.3	212	PASS
174	95	50	100	82.7	62602	PASS
175	174	5	9	7.3	4546	PASS
176	174	95	101	99.4	62221	PASS
177	176	5	9	6.2	3874	PASS