

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052921\
 Data File : VX022254.D
 Acq On : 28 May 2021 15:21
 Operator : JC/MD
 Sample : VSTDCCC020
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC020

Quant Time: May 29 03:50:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X051321W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu May 13 06:11:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.904	128	17318	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.769	114	101542	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	91239	30.00	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.964	65	41930	31.43	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	104.77%
60) 4-Bromofluorobenzene	11.085	95	45858	30.29	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	100.97%
63) Toluene-d8	8.653	98	124287	29.81	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	99.37%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.166	85	20600	20.15	ug/l	96
3) Chloromethane	1.295	50	22661	18.45	ug/l	97
4) Vinyl Chloride	1.374	62	23852	18.70	ug/l	96
5) Bromomethane	1.605	94	13778	19.20	ug/l	100
6) Chloroethane	1.685	64	14922	18.17	ug/l	96
7) Trichlorofluoromethane	1.886	101	34267	19.23	ug/l	99
8) Diethyl Ether	2.142	74	14309	18.72	ug/l	93
9) 1,1,2-Trichlorotrifluo...	2.331	101	19932	19.50	ug/l	99
10) 1,1-Dichloroethene	2.319	96	20017	19.45	ug/l	92
11) Methyl Iodide	2.453	142	21996	21.49	ug/l	98
12) Methyl Acetate	2.709	43	32114	22.56	ug/l	98
13) Acrolein	2.239	56	29089	131.78	ug/l	99
14) Acrylonitrile	3.075	53	63744	109.76	ug/l	97
15) Acetone	2.386	58	20235	116.71	ug/l	87
16) Carbon Disulfide	2.514	76	49033	17.41	ug/l	99
17) Allyl chloride	2.666	41	36617	20.18	ug/l	90
18) Methylene Chloride	2.794	84	23720	19.38	ug/l	96
19) trans-1,2-Dichloroethene	3.093	96	22139	19.79	ug/l	92
20) Diisopropyl ether	3.770	45	70930	20.74	ug/l	95
21) 1,1-Dichloroethane	3.617	63	41060	20.05	ug/l	98
22) cis-1,2-Dichloroethene	4.495	96	26225	19.95	ug/l	97
23) tert-Butyl Alcohol	2.977	59	32069	113.59	ug/l #	100
24) Methyl tert-Butyl Ether	3.123	73	68838	20.54	ug/l	100
25) Chloroform	5.105	83	42324	20.39	ug/l	96
26) Cyclohexane	5.477	56	33924	20.07	ug/l #	99
29) 1,1-Dichloropropene	5.702	75	31211	19.64	ug/l	100
30) 2-Butanone	4.568	43	100449	111.66	ug/l #	97
31) 2,2-Dichloropropane	4.483	77	35638	19.08	ug/l	99
32) 1,1,1-Trichloroethane	5.385	97	35655	19.02	ug/l #	76
33) Carbon Tetrachloride	5.684	117	28753	18.07	ug/l	88
34) Benzene	6.044	78	96012	19.80	ug/l	98
35) Methacrylonitrile	4.934	41	19790	20.55	ug/l	92
36) 1,2-Dichloroethane	6.092	62	33658	20.33	ug/l #	89
37) Trichloroethene	7.135	130	23009	18.88	ug/l	94
38) Methylcyclohexane	7.385	83	34888	19.34	ug/l	98
39) 1,2-Dichloropropane	7.434	63	25126	20.75	ug/l	97
40) Dibromomethane	7.586	93	16275	19.32	ug/l	97
41) Bromodichloromethane	7.824	83	30998	18.43	ug/l	93
42) Vinyl Acetate	3.733	43	341581	101.68	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052921\
 Data File : VX022254.D
 Acq On : 28 May 2021 15:21
 Operator : JC/MD
 Sample : VSTDCCC020
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1


Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC020


Quant Time: May 29 03:50:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X051321W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu May 13 06:11:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
43) Ethyl Acetate	4.733	43	39223	21.46	ug/l	98
44) Isopropyl Acetate	6.348	43	62488	20.73	ug/l	99
45) 1,4-Dioxane	7.665	88	12066	448.24	ug/l #	100
46) Methyl methacrylate	7.696	41	29007	20.75	ug/l	98
47) n-amyl Acetate	10.848	43	52757	21.04	ug/l	100
48) t-1,3-Dichloropropene	8.982	75	35701	18.91	ug/l	95
49) cis-1,3-Dichloropropene	8.372	75	38419	18.96	ug/l	99
50) 1,1,2-Trichloroethane	9.153	97	23710	19.77	ug/l	97
51) Ethyl methacrylate	9.122	69	38870	20.10	ug/l	98
52) 1,3-Dichloropropane	9.311	76	42480	20.32	ug/l	99
53) Dibromochloromethane	9.525	129	20907	17.35	ug/l	98
54) 1,2-Dibromoethane	9.610	107	24032	19.52	ug/l	98
55) 2-Chloroethyl vinyl ether	8.244	63	99211	97.20	ug/l	99
56) Bromoform	10.805	173	12628	16.52	ug/l	97
58) 4-Methyl-2-Pentanone	8.580	43	195526	106.27	ug/l	98
59) 2-Hexanone	9.433	43	147763	108.44	ug/l	99
61) Tetrachloroethene	9.281	164	20249	19.56	ug/l	92
62) Toluene	8.720	91	101045	19.34	ug/l	100
64) Chlorobenzene	10.086	112	60436	19.35	ug/l	98
65) 1,1,1,2-Tetrachloroethane	10.165	131	20376	17.87	ug/l	98
66) Ethyl Benzene	10.195	91	112702	19.61	ug/l	100
67) m/p-Xylenes	10.305	106	84558	38.82	ug/l	98
68) o-Xylene	10.646	106	41828	19.72	ug/l	97
69) Styrene	10.659	104	68220	19.22	ug/l	99
70) Isopropylbenzene	10.963	105	110429	19.82	ug/l	97
71) 1,1,2,2-Tetrachloroethane	11.213	83	37616	20.48	ug/l	98
72) 1,2,3-Trichloropropane	11.244	75	42176	25.90	ug/l	66
73) Bromobenzene	11.201	156	22899	19.40	ug/l	95
74) n-propylbenzene	11.305	91	127163	19.58	ug/l	99
75) 2-Chlorotoluene	11.366	91	75865	19.72	ug/l	100
76) 1,3,5-Trimethylbenzene	11.451	105	91088	19.58	ug/l	100
77) t-1,4-Dichloro-2-butene	11.018	75	11245	17.04	ug/l	76
78) 4-Chlorotoluene	11.457	91	86952	19.64	ug/l	99
79) tert-butylbenzene	11.719	119	85003	19.10	ug/l	99
80) 1,2,4-Trimethylbenzene	11.756	105	90026	19.36	ug/l	98
81) sec-Butylbenzene	11.896	105	110458	19.75	ug/l	98
82) p-Isopropyltoluene	12.012	119	91014	19.41	ug/l	100
83) 1,3-Dichlorobenzene	11.976	146	42893	19.17	ug/l	98
84) 1,4-Dichlorobenzene	12.043	146	42618	18.99	ug/l	98
85) n-Butylbenzene	12.335	91	85455	19.89	ug/l	97
86) Hexachloroethane	12.542	117	12643	16.23	ug/l	94
87) 1,2-Dichlorobenzene	12.335	146	42769	19.73	ug/l	97
88) 1,2-Dibromo-3-Chloropr...	12.945	75	7298	17.84	ug/l	95
89) 1,2,4-Trichlorobenzene	13.591	180	24553	18.49	ug/l	95
90) Hexachlorobutadiene	13.731	225	9657	18.80	ug/l	96
91) Naphthalene	13.780	128	107528	20.64	ug/l	97
92) 1,2,3-Trichlorobenzene	13.963	180	25143	19.07	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: May 29 03:50:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X051321W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu May 13 06:11:35 2021
Response via : Initial Calibration









#1
 Bromochloromethane
 Concen: 30.00 ug/l
 RT: 4.904 min Scan# 626
 Delta R.T. -0.005 min
 Lab File: VX022254.D
 Acq: 28 May 2021 15:21




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC020


Tgt Ion:	128	Resp:	17318
Ion	Ratio	Lower	Upper
128	100		
49	184.1	0.0	442.5
130	128.3	0.0	323.3



#2
 Dichlorodifluoromethane
 Concen: 20.15 ug/l
 RT: 1.166 min Scan# 13
 Delta R.T. 0.000 min
 Lab File: VX022254.D
 Acq: 28 May 2021 15:21

Tgt Ion:	85	Resp:	20600
Ion	Ratio	Lower	Upper
85	100		
87	31.9	14.9	44.5





#3

Chloromethane

Concen: 18.45 ug/l

RT: 1.295 min Scan# 34

Delta R.T. 0.000 min





Lab File: VX022254.D

Acq: 28 May 2021 15:21

Tgt Ion: 50 Resp: 22661

Ion	Ratio	Lower	Upper
50	100		
52	31.4	26.2	39.4

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC020



#4

Vinyl Chloride

Concen: 18.70 ug/l

RT: 1.374 min Scan# 47




Delta R.T. 0.000 min


Lab File: VX022254.D

Acq: 28 May 2021 15:21

Tgt Ion: 62 Resp: 23852

Ion	Ratio	Lower	Upper
62	100		
64	33.3	24.9	37.3









#5
Bromomethane
Concen: 19.20 ug/l
RT: 1.605 min Scan# 85
Delta R.T. -0.006 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC020


Tgt Ion: 94 Resp: 13778
Ion Ratio Lower Upper
94 100
96 95.4 76.1 114.1



#6
Chloroethane
Concen: 18.17 ug/l
RT: 1.685 min Scan# 98
Delta R.T. -0.006 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21

Tgt Ion: 64 Resp: 14922
Ion Ratio Lower Upper
64 100
66 28.7 24.9 37.3









#7
 Trichlorofluoromethane
 Concen: 19.23 ug/l
 RT: 1.886 min Scan# 131
 Delta R.T. -0.006 min
 Lab File: VX022254.D
 Acq: 28 May 2021 15:21




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC020


Tgt Ion:101 Resp: 34267
 Ion Ratio Lower Upper
 101 100
 103 66.7 52.4 78.6



#8
 Diethyl Ether
 Concen: 18.72 ug/l
 RT: 2.142 min Scan# 173
 Delta R.T. 0.007 min
 Lab File: VX022254.D
 Acq: 28 May 2021 15:21

Tgt Ion: 74 Resp: 14309
 Ion Ratio Lower Upper
 74 100
 45 97.9 21.1 189.9









#9
1,1,2-Trichlorotrifluoroethane
Concen: 19.50 ug/l
RT: 2.331 min Scan# 204
Delta R.T. 0.000 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC020


Tgt Ion:101 Resp: 19932
Ion Ratio Lower Upper
101 100
85 48.0 0.0 93.0
151 68.0 0.0 136.6



#10
1,1-Dichloroethene
Concen: 19.45 ug/l
RT: 2.319 min Scan# 202
Delta R.T. -0.006 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21


Tgt Ion: 96 Resp: 20017
Ion Ratio Lower Upper
96 100
61 164.2 121.0 181.4
98 62.2 49.1 73.7








#11
Methyl Iodide
Concen: 21.49 ug/l
RT: 2.453 min Scan# 224
Delta R.T. -0.006 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC020







Tgt Ion:142 Resp: 21996
Ion Ratio Lower Upper
142 100
127 43.0 21.3 63.7
141 12.5 7.2 21.6



#12
Methyl Acetate
Concen: 22.56 ug/l
RT: 2.709 min Scan# 266
Delta R.T. 0.000 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21

Tgt Ion: 43 Resp: 32114
Ion Ratio Lower Upper
43 100
74 24.2 18.5 27.7









#13
Acrolein
Concen: 131.78 ug/l
RT: 2.239 min Scan# 189
Delta R.T. 0.000 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC020


Tgt Ion: 56 Resp: 29089
Ion Ratio Lower Upper
56 100
55 71.0 57.6 86.4



#14
Acrylonitrile
Concen: 109.76 ug/l
RT: 3.075 min Scan# 326
Delta R.T. 0.007 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21

Tgt Ion: 53 Resp: 63744
Ion Ratio Lower Upper
53 100
52 81.1 39.3 117.9
51 35.5 17.1 51.2









#15
Acetone
Concen: 116.71 ug/l
RT: 2.386 min Scan# 213
Delta R.T. 0.001 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC020


Tgt Ion: 58 Resp: 20235
Ion Ratio Lower Upper
58 100
43 286.7 249.5 374.3



#16
Carbon Disulfide
Concen: 17.41 ug/l
RT: 2.514 min Scan# 234
Delta R.T. 0.000 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21

Tgt Ion: 76 Resp: 49033
Ion Ratio Lower Upper
76 100
78 8.9 6.7 10.1





#17

Allyl chloride

Concen: 20.18 ug/l

RT: 2.666 min Scan# 259





Delta R.T. -0.006 min

Lab File: VX022254.D

Acq: 28 May 2021 15:21

Tgt Ion:	41	Resp:	36617
Ion Ratio	Lower	Upper	
41	100		
39	80.0	34.5	103.5
76	41.7	19.8	59.3

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC020



#18

Methylene Chloride

Concen: 19.38 ug/l




RT: 2.794 min Scan# 280


Delta R.T. 0.000 min

Lab File: VX022254.D

Acq: 28 May 2021 15:21

Tgt Ion:	84	Resp:	23720
Ion Ratio	Lower	Upper	
84	100		
49	131.7	101.6	152.4
51	39.8	30.5	45.7
86	63.0	53.4	80.2









#19

trans-1,2-Dichloroethene
Concen: 19.79 ug/l
RT: 3.093 min Scan# 329
Delta R.T. -0.006 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC020




Tgt Ion: 96 Resp: 22139
Ion Ratio Lower Upper
96 100
61 130.1 113.0 169.4
98 59.8 51.7 77.5




#20

Diisopropyl ether
Concen: 20.74 ug/l
RT: 3.770 min Scan# 440
Delta R.T. 0.000 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21

Tgt Ion: 45 Resp: 70930
Ion Ratio Lower Upper
45 100
43 78.7 58.5 87.7
87 28.1 22.2 33.4
59 12.2 9.0 13.6









#21
1,1-Dichloroethane
Concen: 20.05 ug/l
RT: 3.617 min Scan# 415
Delta R.T. 0.000 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC020


Tgt Ion: 63 Resp: 41060
Ion Ratio Lower Upper
63 100
98 7.1 3.8 11.4
100 5.0 2.9 8.6



#22
cis-1,2-Dichloroethene
Concen: 19.95 ug/l
RT: 4.495 min Scan# 559
Delta R.T. -0.006 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21

Tgt Ion: 96 Resp: 26225
Ion Ratio Lower Upper
96 100
61 145.8 114.0 171.0
98 64.2 53.0 79.6









#23

tert-Butyl Alcohol
Concen: 113.59 ug/l
RT: 2.977 min Scan# 310
Delta R.T. 0.006 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC020




Tgt Ion: 59 Resp: 32069
Ion Ratio Lower Upper
59 100
52 0.0 0.0 0.0




#24

Methyl tert-Butyl Ether
Concen: 20.54 ug/l
RT: 3.123 min Scan# 334
Delta R.T. 0.007 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21

Tgt Ion: 73 Resp: 68838
Ion Ratio Lower Upper
73 100
43 20.4 16.3 24.5









#25
Chloroform
Concen: 20.39 ug/l
RT: 5.105 min Scan# 659
Delta R.T. 0.000 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC020


Tgt Ion: 83 Resp: 42324
Ion Ratio Lower Upper
83 100
85 61.4 51.9 77.9



#26
Cyclohexane
Concen: 20.07 ug/l
RT: 5.477 min Scan# 720
Delta R.T. 0.000 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21

Tgt Ion: 56 Resp: 33924
Ion Ratio Lower Upper
56 100
89 1.8 0.4 0.6#
84 89.1 71.8 107.6





#27

1,2-Dichloroethane-d4

Concen: 31.43 ug/l

RT: 5.964 min Scan# 800

Delta R.T. 0.000 min

Lab File: VX022254.D

Acq: 28 May 2021 15:21

Tgt Ion: 65 Resp: 41930

Ion Ratio Lower Upper

65 100





67 53.6 41.5 62.3

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC020



#28

1,4-Difluorobenzene

Concen: 30.00 ug/l

RT: 6.769 min Scan# 932

Delta R.T. 0.000 min

Lab File: VX022254.D

Acq: 28 May 2021 15:21




Tgt Ion:114 Resp: 101542


Ion Ratio Lower Upper

114 100

63 21.2 16.6 25.0

88 17.4 13.1 19.7





#29

1,1-Dichloropropene

Concen: 19.64 ug/l

RT: 5.702 min Scan# 757

Delta R.T. 0.000 min

Lab File: VX022254.D

Acq: 28 May 2021 15:21

Tgt Ion: 75 Resp: 31211

Ion Ratio Lower Upper

75 100

110 36.2 0.0 71.4





77 31.8 0.0 63.6

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC020



#30

2-Butanone

Concen: 111.66 ug/l

RT: 4.568 min Scan# 571

Delta R.T. 0.006 min

Lab File: VX022254.D

Acq: 28 May 2021 15:21




Tgt Ion: 43 Resp: 100449


Ion Ratio Lower Upper

43 100

57 4.1 6.4 9.6#

72 26.9 22.1 33.1









#31
2,2-Dichloropropane
Concen: 19.08 ug/l
RT: 4.483 min Scan# 557
Delta R.T. -0.006 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC020


Tgt Ion: 77 Resp: 35638
Ion Ratio Lower Upper
77 100
97 23.1 18.1 27.1



#32
1,1,1-Trichloroethane
Concen: 19.02 ug/l
RT: 5.385 min Scan# 705
Delta R.T. -0.006 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21

Tgt Ion: 97 Resp: 35655
Ion Ratio Lower Upper
97 100
99 33.3 51.9 77.9#
61 45.4 36.3 54.5









#33
Carbon Tetrachloride
Concen: 18.07 ug/l
RT: 5.684 min Scan# 754
Delta R.T. -0.006 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC020


Tgt Ion:117 Resp: 28753
Ion Ratio Lower Upper
117 100
119 86.8 80.8 121.2
121 26.4 23.8 35.8



#34
Benzene
Concen: 19.80 ug/l
RT: 6.044 min Scan# 813
Delta R.T. 0.000 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21

Tgt Ion: 78 Resp: 96012
Ion Ratio Lower Upper
78 100
77 22.9 17.8 26.8
51 15.1 10.6 15.8





#35

Methacrylonitrile

Concen: 20.55 ug/l

RT: 4.934 min Scan# 631

Delta R.T. 0.007 min

Lab File: VX022254.D

Acq: 28 May 2021 15:21

Tgt Ion: 41 Resp: 19790





Ion Ratio Lower Upper

41 100

39 57.0 29.3 87.8

67 64.7 38.3 114.9

52 29.2 15.9 47.6



#36

1,2-Dichloroethane

Concen: 20.33 ug/l

RT: 6.092 min Scan# 821

Delta R.T. 0.000 min

Lab File: VX022254.D




Acq: 28 May 2021 15:21


Tgt Ion: 62 Resp: 33658

Ion Ratio Lower Upper

62 100

98 9.6 4.6 7.0#





#37

Trichloroethene

Concen: 18.88 ug/l

RT: 7.135 min Scan# 992

Delta R.T. 0.000 min





Lab File: VX022254.D

Acq: 28 May 2021 15:21

Tgt Ion:130 Resp: 23009

Ion Ratio Lower Upper

130	100		
95	107.5	81.1	121.7



#38

Methylcyclohexane

Concen: 19.34 ug/l

RT: 7.385 min Scan# 1033

Delta R.T. 0.000 min




Lab File: VX022254.D


Acq: 28 May 2021 15:21

Tgt Ion: 83 Resp: 34888

Ion Ratio Lower Upper

83	100		
55	75.8	36.4	109.2
98	50.7	25.3	75.9









#39
1,2-Dichloropropane
Concen: 20.75 ug/l
RT: 7.434 min Scan# 1041
Delta R.T. -0.006 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC020


Tgt Ion: 63 Resp: 25126
Ion Ratio Lower Upper
63 100
65 33.1 25.2 37.8



#40
Dibromomethane
Concen: 19.32 ug/l
RT: 7.586 min Scan# 1066
Delta R.T. 0.000 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21

Tgt Ion: 93 Resp: 16275
Ion Ratio Lower Upper
93 100
95 82.5 0.0 166.0
174 80.7 0.0 170.4









#41
Bromodichloromethane
Concen: 18.43 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. -0.006 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC020


Tgt Ion:	83	Resp:	30998
Ion	Ratio	Lower	Upper
83	100		
85	59.4	52.6	78.8
127	7.8	6.1	9.1



#42
Vinyl Acetate
Concen: 101.68 ug/l
RT: 3.733 min Scan# 434
Delta R.T. 0.006 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21

Tgt Ion:	43	Resp:	341581
Ion	Ratio	Lower	Upper
43	100		
86	9.7	8.2	12.4









#43
Ethyl Acetate
Concen: 21.46 ug/l
RT: 4.733 min Scan# 598
Delta R.T. 0.013 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC020


Tgt Ion: 43 Resp: 39223
Ion Ratio Lower Upper
43 100
45 14.6 11.2 16.8



#44
Isopropyl Acetate
Concen: 20.73 ug/l
RT: 6.348 min Scan# 863
Delta R.T. 0.000 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21

Tgt Ion: 43 Resp: 62488
Ion Ratio Lower Upper
43 100
61 20.5 17.0 25.4









#45
1,4-Dioxane
Concen: 448.24 ug/l
RT: 7.665 min Scan# 1079
Delta R.T. -0.012 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC020


Tgt Ion: 88 Resp: 12066
Ion Ratio Lower Upper
88 100
43 36.2 0.0 0.0#
58 65.5 0.0 0.0#



#46
Methyl methacrylate
Concen: 20.75 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. 0.000 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21

Tgt Ion: 41 Resp: 29007
Ion Ratio Lower Upper
41 100
69 85.3 42.7 128.1
39 60.7 32.1 96.3









#47
n-amyl Acetate
Concen: 21.04 ug/l
RT: 10.848 min Scan# 1601
Delta R.T. 0.000 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC020


Tgt Ion: 43 Resp: 52757
Ion Ratio Lower Upper
43 100
55 23.5 18.9 28.3



#48
t-1,3-Dichloropropene
Concen: 18.91 ug/l
RT: 8.982 min Scan# 1295
Delta R.T. 0.000 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21

Tgt Ion: 75 Resp: 35701
Ion Ratio Lower Upper
75 100
77 30.9 26.8 40.2









#49
cis-1,3-Dichloropropene
Concen: 18.96 ug/l
RT: 8.372 min Scan# 1195
Delta R.T. 0.000 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC020


Tgt Ion:	75	Resp:	38419
Ion	Ratio	Lower	Upper
75	100		
77	31.7	24.9	37.3
39	49.1	38.6	57.8



#50
1,1,2-Trichloroethane
Concen: 19.77 ug/l
RT: 9.153 min Scan# 1323
Delta R.T. -0.006 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21

Tgt Ion:	97	Resp:	23710
Ion	Ratio	Lower	Upper
97	100		
83	87.9	68.2	102.4
85	56.6	44.4	66.6
99	59.9	44.2	66.4









#51
Ethyl methacrylate
Concen: 20.10 ug/l
RT: 9.122 min Scan# 1318
Delta R.T. 0.000 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC020


Tgt Ion: 69 Resp: 38870
Ion Ratio Lower Upper
69 100
41 61.4 31.1 93.5
39 35.2 18.3 54.9



#52
1,3-Dichloropropane
Concen: 20.32 ug/l
RT: 9.311 min Scan# 1349
Delta R.T. 0.000 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21

Tgt Ion: 76 Resp: 42480
Ion Ratio Lower Upper
76 100
78 31.7 0.0 64.2





#53

Dibromochloromethane

Concen: 17.35 ug/l

RT: 9.525 min Scan# 1384

Delta R.T. 0.000 min

Lab File: VX022254.D

Acq: 28 May 2021 15:21

Tgt Ion:129 Resp: 20907

Ion Ratio Lower Upper

129 100





127 76.4 39.0 116.9

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC020



#54

1,2-Dibromoethane

Concen: 19.52 ug/l

RT: 9.610 min Scan# 1398

Delta R.T. 0.000 min

Lab File: VX022254.D




Acq: 28 May 2021 15:21


Tgt Ion:107 Resp: 24032

Ion Ratio Lower Upper

107 100

109 95.2 74.4 111.6









#55
2-Chloroethyl vinyl ether
Concen: 97.20 ug/l
RT: 8.244 min Scan# 1174
Delta R.T. 0.000 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC020


Tgt Ion: 63 Resp: 99211
Ion Ratio Lower Upper
63 100
65 32.0 25.6 38.4
106 27.6 23.2 34.8



#56
Bromoform
Concen: 16.52 ug/l
RT: 10.805 min Scan# 1594
Delta R.T. 0.000 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21

Tgt Ion: 173 Resp: 12628
Ion Ratio Lower Upper
173 100
175 47.2 39.4 59.2
254 8.1 8.0 12.0









#57
Chlorobenzene-d5
Concen: 30.00 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC020


Tgt Ion:117 Resp: 91239
Ion Ratio Lower Upper
117 100
82 57.7 44.9 67.3
119 31.0 25.0 37.6



#58
4-Methyl-2-Pentanone
Concen: 106.27 ug/l
RT: 8.580 min Scan# 1229
Delta R.T. 0.001 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21

Tgt Ion: 43 Resp: 195526
Ion Ratio Lower Upper
43 100
58 38.9 31.8 47.6
85 17.5 14.9 22.3









#59
2-Hexanone
Concen: 108.44 ug/l
RT: 9.433 min Scan# 1369
Delta R.T. 0.000 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC020


Tgt Ion:	43	Resp:	147763
Ion	Ratio	Lower	Upper
43	100		
58	55.3	45.3	67.9
57	18.5	15.0	22.6



#60
4-Bromofluorobenzene
Concen: 30.29 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. 0.000 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21


Tgt Ion:	95	Resp:	45858
Ion	Ratio	Lower	Upper
95	100		
174	64.3	52.2	78.2
176	61.0	49.3	73.9








#61
Tetrachloroethene
Concen: 19.56 ug/l
RT: 9.281 min Scan# 1344
Delta R.T. 0.001 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC020







Tgt Ion:	164	Resp:	20249
Ion Ratio	Lower	Upper	
164	100		
166	124.6	90.2	135.4
129	96.9	73.4	110.0
131	87.2	75.4	113.0



#62
Toluene
Concen: 19.34 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. 0.000 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21

Tgt Ion:	91	Resp:	101045
Ion Ratio	Lower	Upper	
91	100		
92	59.1	47.4	71.2









#63
Toluene-d8
Concen: 29.81 ug/l
RT: 8.653 min Scan# 1241
Delta R.T. 0.001 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC020


Tgt Ion: 98 Resp: 124287
Ion Ratio Lower Upper
98 100
100 66.0 52.6 78.8



#64
Chlorobenzene
Concen: 19.35 ug/l
RT: 10.086 min Scan# 1476
Delta R.T. 0.000 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21

Tgt Ion: 112 Resp: 60436
Ion Ratio Lower Upper
112 100
114 30.5 25.3 37.9









#65
1,1,1,2-Tetrachloroethane
Concen: 17.87 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. 0.000 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC020


Tgt Ion:131 Resp: 20376
Ion Ratio Lower Upper
131 100
133 96.4 0.0 187.6
119 68.4 0.0 137.6



#66
Ethyl Benzene
Concen: 19.61 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.000 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21

Tgt Ion: 91 Resp: 112702
Ion Ratio Lower Upper
91 100
106 30.9 24.6 37.0









#67
m/p-Xylenes
Concen: 38.82 ug/l
RT: 10.305 min Scan# 1512
Delta R.T. 0.000 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC020


Tgt Ion:106 Resp: 84558
Ion Ratio Lower Upper
106 100
91 204.2 166.0 249.0



#68
o-Xylene
Concen: 19.72 ug/l
RT: 10.646 min Scan# 1568
Delta R.T. 0.000 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21

Tgt Ion:106 Resp: 41828
Ion Ratio Lower Upper
106 100
91 216.7 106.0 318.0





#69

Styrene

Concen: 19.22 ug/l

RT: 10.659 min Scan# 1570

Delta R.T. 0.000 min

Lab File: VX022254.D

Acq: 28 May 2021 15:21





Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC020

Tgt Ion:104	Resp:	68220
Ion Ratio	Lower	Upper
104	100	
78	51.9	41.8 62.8
103	54.0	44.1 66.1



#70

Isopropylbenzene

Concen: 19.82 ug/l




RT: 10.963 min Scan# 1620


Delta R.T. 0.000 min

Lab File: VX022254.D

Acq: 28 May 2021 15:21

Tgt Ion:105	Resp:	110429
Ion Ratio	Lower	Upper
105	100	
120	25.7	18.9 35.1









#71
1,1,2,2-Tetrachloroethane
Concen: 20.48 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. -0.006 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC020


Tgt Ion: 83 Resp: 37616
Ion Ratio Lower Upper
83 100
131 8.8 4.7 14.1
85 63.4 30.8 92.3



#72
1,2,3-Trichloropropane
Concen: 25.90 ug/l
RT: 11.244 min Scan# 1666
Delta R.T. 0.000 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21

Tgt Ion: 75 Resp: 42176
Ion Ratio Lower Upper
75 100
77 31.1 0.0 33.0









#73
Bromobenzene
Concen: 19.40 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. 0.000 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC020


Tgt Ion:156 Resp: 22899
Ion Ratio Lower Upper
156 100
77 190.8 0.0 363.4
158 93.2 0.0 193.4



#74
n-propylbenzene
Concen: 19.58 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. -0.006 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21

Tgt Ion: 91 Resp: 127163
Ion Ratio Lower Upper
91 100
120 21.4 0.0 44.2









#75
2-Chlorotoluene
Concen: 19.72 ug/l
RT: 11.366 min Scan# 1686
Delta R.T. 0.000 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC020


Tgt Ion: 91 Resp: 75865
Ion Ratio Lower Upper
91 100
126 32.0 0.0 63.8



#76
1,3,5-Trimethylbenzene
Concen: 19.58 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.006 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21

Tgt Ion: 105 Resp: 91088
Ion Ratio Lower Upper
105 100
120 48.0 0.0 96.2









#77
t-1,4-Dichloro-2-butene
Concen: 17.04 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. -0.006 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC020


Tgt Ion: 75 Resp: 11245
Ion Ratio Lower Upper
75 100
53 107.5 40.2 120.6
89 38.2 23.4 70.0



#78
4-Chlorotoluene
Concen: 19.64 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21

Tgt Ion: 91 Resp: 86952
Ion Ratio Lower Upper
91 100
126 28.2 0.0 55.8









#79
tert-butylbenzene
Concen: 19.10 ug/l
RT: 11.719 min Scan# 1744
Delta R.T. 0.000 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC020


Tgt Ion:	119	Resp:	85003
Ion Ratio	Lower	Upper	
119	100		
91	60.9	0.0	123.0
134	23.5	0.0	45.4



#80
1,2,4-Trimethylbenzene
Concen: 19.36 ug/l
RT: 11.756 min Scan# 1750
Delta R.T. 0.000 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21

Tgt Ion:	105	Resp:	90026
Ion Ratio	Lower	Upper	
105	100		
120	44.4	0.0	91.4









#81
 sec-Butylbenzene
 Concen: 19.75 ug/l
 RT: 11.896 min Scan# 1773
 Delta R.T. 0.000 min
 Lab File: VX022254.D
 Acq: 28 May 2021 15:21




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC020


Tgt Ion:105 Resp: 110458
 Ion Ratio Lower Upper
 105 100
 134 18.3 0.0 38.8



#82
 p-Isopropyltoluene
 Concen: 19.41 ug/l
 RT: 12.012 min Scan# 1792
 Delta R.T. 0.000 min
 Lab File: VX022254.D
 Acq: 28 May 2021 15:21

Tgt Ion:119 Resp: 91014
 Ion Ratio Lower Upper
 119 100
 134 25.4 0.0 50.4
 91 25.6 0.0 51.0









#83
1,3-Dichlorobenzene
Concen: 19.17 ug/l
RT: 11.976 min Scan# 1786
Delta R.T. 0.000 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC020


Tgt Ion:146 Resp: 42893
Ion Ratio Lower Upper
146 100
111 46.6 22.1 66.5
148 63.4 31.1 93.5



#84
1,4-Dichlorobenzene
Concen: 18.99 ug/l
RT: 12.043 min Scan# 1797
Delta R.T. 0.000 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21

Tgt Ion:146 Resp: 42618
Ion Ratio Lower Upper
146 100
111 44.1 21.0 63.0
148 61.9 31.6 95.0









#85
n-Butylbenzene
Concen: 19.89 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC020


Tgt Ion:	91	Resp:	85455
Ion	Ratio	Lower	Upper
91	100		
92	53.3	0.0	111.0
134	24.0	0.0	49.8



#86
Hexachloroethane
Concen: 16.23 ug/l
RT: 12.542 min Scan# 1879
Delta R.T. 0.000 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21

Tgt Ion:	117	Resp:	12643
Ion	Ratio	Lower	Upper
117	100		
201	57.3	30.9	92.8









#87
1,2-Dichlorobenzene
Concen: 19.73 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.006 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC020


Tgt Ion:146 Resp: 42769
Ion Ratio Lower Upper
146 100
111 44.8 22.9 68.8
148 60.2 31.9 95.9



#88
1,2-Dibromo-3-Chloropropane
Concen: 17.84 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. 0.000 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21

Tgt Ion: 75 Resp: 7298
Ion Ratio Lower Upper
75 100
155 74.5 64.6 97.0
157 97.6 75.8 113.6









#89
1,2,4-Trichlorobenzene
Concen: 18.49 ug/l
RT: 13.591 min Scan# 2051
Delta R.T. 0.000 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC020


Tgt Ion:180 Resp: 24553
Ion Ratio Lower Upper
180 100
182 93.6 77.8 116.6
145 33.3 30.0 45.0



#90
Hexachlorobutadiene
Concen: 18.80 ug/l
RT: 13.731 min Scan# 2074
Delta R.T. 0.007 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21

Tgt Ion:225 Resp: 9657
Ion Ratio Lower Upper
225 100
223 61.3 0.0 128.0
227 61.4 0.0 129.6









#91
Naphthalene
Concen: 20.64 ug/l
RT: 13.780 min Scan# 2082
Delta R.T. 0.000 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC020

Tgt Ion:128 Resp: 107528
Ion Ratio Lower Upper
128 100
127 11.9 10.4 15.6
129 10.5 9.2 13.8



#92
1,2,3-Trichlorobenzene
Concen: 19.07 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. 0.000 min
Lab File: VX022254.D
Acq: 28 May 2021 15:21

Tgt Ion:180 Resp: 25143
Ion Ratio Lower Upper
180 100
182 95.1 0.0 192.6
145 36.4 0.0 75.8

