

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052921\
 Data File : VX022250.D
 Acq On : 28 May 2021 13:42
 Operator : JC/MD
 Sample : VX0529MBL01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0529MBL01

Quant Time: May 28 14:04:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	63751	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	131094	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	111203	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	40819	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	52108	52.99	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 105.98%	
35) Dibromofluoromethane	5.391	113	39368	47.36	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 94.72%	
50) Toluene-d8	8.653	98	158186	49.40	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 98.80%	
62) 4-Bromofluorobenzene	11.085	95	52515	44.37	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 88.74%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052921\
Data File : VX022250.D
Acq On : 28 May 2021 13:42
Operator : JC/MD
Sample : VX0529MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0529MBL01

Quant Time: May 28 14:04:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

