

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052921\
 Data File : VX022257.D
 Acq On : 28 May 2021 16:47
 Operator : JC/MD
 Sample : M2483-02DL 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DSN-002-SEWER-OUTLET-002DL

Manual Integrations
 APPROVED

MMDadoda
 6/2/2021 2:27:55 PM

Quant Time: May 29 03:52:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X051321W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu May 13 06:11:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.910	128	17973	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.769	114	103956	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	88806	30.00	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.964	65	42979	31.04	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	103.47%
60) 4-Bromofluorobenzene	11.085	95	43204	29.31	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	97.70%
63) Toluene-d8	8.653	98	125880	31.02	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	103.40%
Target Compounds						
15) Acetone	2.386	58	46514	258.50	ug/l	99
25) Chloroform	5.111	83	3056m	1.42	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052921\
Data File : VX022257.D
Acq On : 28 May 2021 16:47
Operator : JC/MD
Sample : M2483-02DL 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DSN-002-SEWER-OUTLET-002DL

Manual Integrations
APPROVED

MMDadoda
6/2/2021 2:27:55 PM

Quant Time: May 29 03:52:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X051321W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu May 13 06:11:35 2021
Response via : Initial Calibration

