

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052921\
 Data File : VX022275.D
 Acq On : 29 May 2021 01:06
 Operator : JC/MD
 Sample : M2561-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 36BR-20210526

Quant Time: May 29 04:58:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	70472	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.775	114	148621	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	136174	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	56448	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	58229	53.56	ug/l	0.00
Spiked Amount 50.000	Range 78	- 117	Recovery	= 107.12%		
35) Dibromofluoromethane	5.403	113	40739	43.23	ug/l	0.01
Spiked Amount 50.000	Range 75	- 124	Recovery	= 86.46%		
50) Toluene-d8	8.653	98	182825	50.36	ug/l	0.00
Spiked Amount 50.000	Range 92	- 112	Recovery	= 100.72%		
62) 4-Bromofluorobenzene	11.085	95	67010	49.94	ug/l	0.00
Spiked Amount 50.000	Range 83	- 123	Recovery	= 99.88%		
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.172	85	53185	62.08	ug/l	97
4) Vinyl Chloride	1.374	62	5534760	5659.96	ug/l	89
9) 1,1,2-Trichlorotrifluo...	2.331	101	878	1.05	ug/l #	56
12) 1,1-Dichloroethene	2.325	96	174347	215.86	ug/l	97
16) Acetone	2.392	43	5023	10.61	ug/l #	80
17) Carbon Disulfide	2.526	76	49266	23.29	ug/l	98
21) trans-1,2-Dichloroethene	3.105	96	332079	378.00	ug/l	99
27) cis-1,2-Dichloroethene	4.538	96	26188872	24512.11	ug/l	78
29) Tetrahydrofuran	5.031	42	117190	257.12	ug/l	90
40) Benzene	6.056	78	9663	2.29	ug/l	99
44) Trichloroethene	7.190	130	18975580	18058.94	ug/l	93
51) 4-Methyl-2-Pentanone	8.580	43	9189	5.57	ug/l	94
52) Toluene	8.726	92	13557	5.06	ug/l	99
64) Tetrachloroethene	9.281	164	432	0.48	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052921\
Data File : VX022275.D
Acq On : 29 May 2021 01:06
Operator : JC/MD
Sample : M2561-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
36BR-20210526

Quant Time: May 29 04:58:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

