

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052921\
 Data File : VX022279.D
 Acq On : 29 May 2021 02:40
 Operator : JC/MD
 Sample : M2529-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 41822

Quant Time: May 29 04:59:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	62480	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	130514	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	114519	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	42771	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	52361	54.33	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 108.66%		
35) Dibromofluoromethane	5.397	113	39684	47.95	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 95.90%		
50) Toluene-d8	8.653	98	161613	50.69	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 101.38%		
62) 4-Bromofluorobenzene	11.085	95	55392	47.00	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 94.00%		
Target Compounds						
						Qvalue
3) Chloromethane	1.301	50	359	0.44	ug/l	# 86
4) Vinyl Chloride	1.374	62	839	0.97	ug/l	89
16) Acetone	2.392	43	1741	4.15	ug/l	# 75
19) Methyl tert-butyl Ether	3.123	73	2277	0.93	ug/l	93
27) cis-1,2-Dichloroethene	4.501	96	6292	6.64	ug/l	93
44) Trichloroethene	7.135	130	16008	17.35	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052921\
Data File : VX022279.D
Acq On : 29 May 2021 02:40
Operator : JC/MD
Sample : M2529-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
41822

Quant Time: May 29 04:59:51 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

