

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052921\
 Data File : VX022281.D
 Acq On : 29 May 2021 03:27
 Operator : JC/MD
 Sample : PB136709TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB136709TB

Quant Time: May 29 05:00:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	64245	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	138745	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	127908	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	56484	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	55343	55.84	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 111.68%	
35) Dibromofluoromethane	5.391	113	41661	47.35	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 94.70%	
50) Toluene-d8	8.653	98	171670	50.65	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 101.30%	
62) 4-Bromofluorobenzene	11.085	95	65216	52.06	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 104.12%	
Target Compounds						
						Qvalue
16) Acetone	2.385	43	19119	44.31	ug/l	91
20) Methylene Chloride	2.788	84	3639	3.80	ug/l	94
27) cis-1,2-Dichloroethene	4.495	96	3206	3.29	ug/l	90
43) Isopropyl Acetate	6.348	43	15015	5.84	ug/l	93
44) Trichloroethene	7.135	130	7335	7.48	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052921\
Data File : VX022281.D
Acq On : 29 May 2021 03:27
Operator : JC/MD
Sample : PB136709TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB136709TB

Quant Time: May 29 05:00:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

